

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005940 Location: ToxL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 451 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.030 AC
Sim. Ser #: MP5289 Lot #: 201911E Exp. Date: 11-19-21
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3062 Lot #: 202103E Exp. Date: 3-24-23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 26 Exp. Date: 3-5-23

Test 1	<u>0.079</u>	Test 4	<u>0.078</u>	Test 7	<u>0.078</u>
Test 2	<u>0.078</u>	Test 5	<u>0.078</u>	Test 8	<u>0.078</u>
Test 3	<u>0.078</u>	Test 6	<u>0.078</u>	Test 9	<u>0.079</u>
Average <u>0.078</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED SIMULATOR RETURN O-RING

Instrument is acceptable to be used in the field.

Charles E. Ed
Breath Analyst Signature

6/11/21
Date

Kali J. Dist
Reviewed by

6-14-21
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:20
02 Std. Sol.	0.028	15:21
03 Room Air	0.000	15:21
04 Std. Sol.	0.028	15:22
05 Room Air	0.000	15:23
06 Std. Sol.	0.028	15:23
07 Room Air	0.000	15:24

08 Sim Temp = 34.0°C

Simul Ser No = MP5289
Std Sol No = 201911E
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low Ac*
0.030Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:30
02 Std. Sol.	0.400	15:31
03 Room Air	0.000	15:32
04 Std. Sol.	0.401	15:32
05 Room Air	0.000	15:33
06 Std. Sol.	0.401	15:34
07 Room Air	0.000	15:34


08 Sim Temp = 34.0°C

Simul Ser No = MP3062

Std Sol No = 202103E

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:36

Test	AC	Time
01 Room Air	0.000	15:37
02 *Subject Test	INT*	15:37
03 Room Air	0.000	15:38

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 26
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE
cee
0.05%

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:38

Test	AC	Time
01 Diagnostic	OK	15:39
02 Room Air	0.000	15:40
03 *Subject Test	RFI*	15:40
04 Room Air	0.000	15:40

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 26
Expiration Date = 03/05/2023
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:41
02 Std. Gas	0.079	15:41
03 Room Air	0.000	15:42
04 Std. Gas	0.078	15:42
05 Room Air	0.000	15:43
06 Std. Gas	0.078	15:43
07 Room Air	0.000	15:44

Lot No = 02621080A1
Cyl No = 26
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:44
02 Std. Gas	0.078	15:45
03 Room Air	0.000	15:45
04 Std. Gas	0.078	15:46
05 Room Air	0.000	15:46
06 Std. Gas	0.078	15:46
07 Room Air	0.000	15:47

Lot No = 02621080A1
Cyl No = 26
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/11/2021 15:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:48
02 Std. Gas	0.078	15:48
03 Room Air	0.000	15:49
04 Std. Gas	0.078	15:49
05 Room Air	0.000	15:49
06 Std. Gas	0.079	15:50
07 Room Air	0.000	15:50

Lot No = 02621080A1
Cyl No = 26
Exp Date = 03/05/2023
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000