



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|---|---|
| Serial Number 80-006500 | Instrument Location Harvey PD |
| Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____ | |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | |
|-------------------------------|--------------------------|
| Field Inspector Signature | Date 10-Jul-21 |
|-------------------------------|--------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|---------------------------|------------------------------------|
| Reviewed/Certified By | Certified Date 7.20.2021 |
|---------------------------|------------------------------------|

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.14.00 09/16
07/10/2021 16:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 14
Standard Expiration? 03/05/2023
Oper No? 131495

Flow Cal. Date: 08/18/2015
Slope 681
Intercept -680985

IR Calibration Date: 08/18/2015
3um 9um

0th Coef(*100): -3148 -17259
1st Coef(*100): 254475 138520
2nd Coef(*100): 1954 1329
H2O adj(mg/l*10k): 193 414

***** Printer Test End *****



Operator Signature
MICHAEL BEAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.14.00 09/16
07/10/2021 16:24

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 16:25 |
| 02 Std. Gas | 0.079 | 16:25 |
| 03 Room Air | 0.000 | 16:25 |
| 04 Std. Gas | 0.079 | 16:26 |
| 05 Room Air | 0.000 | 16:26 |
| 06 Std. Gas | 0.079 | 16:27 |
| 07 Room Air | 0.000 | 16:27 |

Lot No = 02621080A1
Cyl No = 14
Exp Date = 03/05/2023
County = 52 Oper No. = 131495



Operator Signature
MICHAEL BEAVER

Remarks: *OK*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.14.00 09/16
07/10/2021 16:30

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | RFI* | 16:32 |
| 02 Room Air | 0.000 | 16:33 |

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 02/02/1990
Sub Sex = Male Weight = 100
Test = DUI Cit = 0000
Dr. Lic. = ND/00000
Lot No = 02621080A1
Cyl No = 14
Expiration Date = 03/05/2023
County = 52 Oper No. = 131495



Operator Signature
MICHAEL BEAVER

Remarks:



Form 106-I8000