



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006504	Instrument Location WCSSO
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Install after expired FI installed	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **N/A**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

CEE
8.17.21

REC'D RFI TEST

CEE
8.17.21

REC'D ALL TEST RECORDS

Field Inspector Signature	Date 17-Aug-21
---------------------------	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	Certified Date 8/17/2021
-----------------------	---------------------------------


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = WESO 8164.14.00 09/16
08/17/2021 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.079	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.078	10:20
05 Room Air	0.000	10:20
06 Std. Gas	0.079	10:20
07 Room Air	0.000	10:21

Lot No = 24119080A1
Cyl No = 12
Exp Date = 11/05/2021
County = 52 Oper No. = 131495


Operator Signature
MICHAEL BEAVER

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = WESO 8164.14.00 09/16
08/17/2021 10:25

Test	AC	Time
01 Room Air	RFI*	10:30
02 Room Air	0.000	10:30

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/01/2000
Sub Sex = Male Weight = 150
Test = DUI Cit = 0000
Dr. Lic. = ND/000
Lot No = 24119080A1
Cyl No = 12
Expiration Date = 11/05/2021
County = 52 Oper No. = 131495



Operator Signature
MICHAEL BEAVER

Remarks:

Form 106-I8000

★ Wells County Sheriff's Office ★

Christopher Kluth, Sheriff
700 Railway St N – Suite 306
Fessenden, ND 58438
701-547-3211 (office)
701-547-2536 (fax)

INSTALL LOOKS GOOD
BUT ANDREW GIVENS
EXPIRED AS FIELD INSP.
ON 2/1/2020
CEE 5.27.21

Date: 5-27-2021

CALLED ON 5.27.21 AND
TALKED TO SHERIFF
CEE 5.27.21

To: Charles

Email/Fax Number: 701-328-6185

EMAILED AGENCY (GIVENS
AND BEAVER) ON 7.29.21
CEE 7.29.21

Re:

- From:
- Christopher Kluth, Sheriff chrkluth@nd.gov
 - Andrew Givens, Chief Deputy angivens@nd.gov
 - Michael Beaver, Deputy mpbeaver@nd.gov
 - Howard Longie, Deputy holongie@nd.gov
 - Janelle Pepple, Office Deputy jpepple@nd.gov

Number of Pages (including cover): 5

Notes: Intox 8000 install

This transmission is intended exclusively for the individual or entity to which it is addressed and may contain information that is confidential and/or privileged. Any unauthorized review, use, copying, disclosure or distribution is strongly prohibited. If you have received this communication in error, please immediately advise the sender by telephone and properly destroy the document.



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number 80-006504	Instrument Location Wells County S.O. Fessenden ND
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 5-27-2021
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	Certified Date
-----------------------	----------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = WESO 8164.14.00 09/16
05/26/2021 10:55

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 12
Standard Expiration? 11/05/2021
Oper No? 132065

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 05/31/2017

	3um	9um
0th Coef(*100):	-21488	-16160
1st Coef(*100):	265581	137324
2nd Coef(*100):	2507	1149
H2O adj(mg/l*10k):	519	348

***** Printer Test End *****



Operator Signature
ANDREW GIVENS

Remarks:


Intoxilyzer Test Record and Checklist.
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = WESO 8164.14.00 09/16
05/26/2021 10:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Gas	0.080	10:59
03 Room Air	0.000	11:00
04 Std. Gas	0.081	11:00
05 Room Air	0.000	11:01
06 Std. Gas	0.081	11:01
07 Room Air	0.000	11:02

Lot No = 24119080A1
Cyl No = 12
Exp Date = 11/05/2021
County = 52 Oper No. = 132065



Operator Signature
ANDREW GIVENS

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = WESO 8164.14.00 09/16
05/27/2021 09:27

Test	AC	Time
01 Room Air	0.000	09:30
02 *Subject Test	RFI*	09:30
03 Room Air	0.000	09:31

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/30/1995
Sub Sex = Male Weight = 160
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 24119080A1
Cyl No = 12
Expiration Date = 11/05/2021
County = 52 Oper No. = 132065



Operator Signature
ANDREW GIVENS

Remarks:

Form 106-I8000