



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <u>Jared Lemieux</u>		
Location <u>BMDC</u>	Intoxilyzer® Serial Number <u>80-004945</u>	
Gas Lot Number <u>02621080A1</u>	Gas Cylinder Number <u>7</u>	Gas Expiration Date <u>3/5/2023</u>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.080 AC
 - B. 0.080 AC
 - C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <u>[Signature]</u> 4834	Date <u>10/1/2021</u>
Reviewed By (Crime Laboratory Use Only) <u>[Signature]</u>	Date <u>10/1/2021</u>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = BMDC 8164.14.00 09/16
10/01/2021 05:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	05:53
02 Std. Gas	0.080	05:54
03 Room Air	0.000	05:54
04 Std. Gas	0.080	05:54
05 Room Air	0.000	05:55
06 Std. Gas	0.080	05:55
07 Room Air	0.000	05:56

Lot No = 02621080A1
Cyl No = 7
Exp Date = 03/05/2023
County = 08 Oper No. = 130315


Operator Signature
JARED LEMIEUX

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = BMDC 8164.14.00 09/16
10/01/2021 05:52

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 7
Standard Expiration? 03/05/2023
Oper No? 130315

Flow Cal. Date: 09/26/2012
Slope 622
Intercept -485836

IR Calibration Date: 04/22/2016
 3um 9um

0th Coef(*100): -23427 -26406
1st Coef(*100): 266989 144341
2nd Coef(*100): 2085 1078
H2O adj(mg/l*10k): 539 400

***** Printer Test End *****



Operator Signature
JARED LEMIEUX

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = BMDC 8164.14.00 09/16
10/01/2021 05:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	05:57
02 Room Air	0.000	05:57

*Invalid Test
Inhibited - RFI

Lot No = 02621080A1
Cyl No = 7
Exp Date = 03/05/2023
County = 08 Oper No. = 130315


Operator Signature
JARED LEMIEUX

Remarks:

Form 106-I8000