



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) JACOB OLSON 42317893		
Location MINOT POLICE DEPT.	Intoxilyzer® Serial Number 80-006503	
Gas Lot Number 02621080A1	Gas Cylinder Number 10	Gas Expiration Date 03-05-2023

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. **0.021** AC
 B. **0.021** AC
 C. **0.021** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>Jacob Olson 42317893</i>	Date 10-02-2021
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ehr</i>	Date 10/4/2021

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = MIPD 8164.14.00 09/16
10/02/2021 16:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:03
02 Std. Gas	0.081	16:03
03 Room Air	0.000	16:04
04 Std. Gas	0.081	16:04
05 Room Air	0.000	16:05
06 Std. Gas	0.081	16:05
07 Room Air	0.000	16:06

Lot No = 02621080A1
Cyl No = 10
Exp Date = 03/05/2023
County = 51 Oper No. = 130907


Operator Signature
JACOB OLSON

Remarks:

Form 106-I8000