



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Holm, Shawn</i>		
Location <i>Divide County So Intake/Processing</i>		Intoxilyzer® Serial Number <i>80-004184</i>
Gas Lot Number <i>26021080A1</i>	Gas Cylinder Number <i>24</i>	Gas Expiration Date <i>10/5/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.082* AC  
 B. *0.082* AC  
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:  
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).  
 B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>11-8-21</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>11/16/2021</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004184  
Location = CROS      8164.14.00 09/16  
11/08/2021      22:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:49
02 Std. Gas	0.082	22:49
03 Room Air	0.000	22:50
04 Std. Gas	0.082	22:50
05 Room Air	0.000	22:51
06 Std. Gas	0.082	22:51
07 Room Air	0.000	22:52

Lot No = 26021080A1  
Cyl No = 24  
Exp Date = 10/05/2023  
County = 12      Oper No. = 131836

  
\_\_\_\_\_  
Operator Signature  
SHAWN HOLM

Remarks:

The undersigned, having custody of the original record, certifies that the information herein is a true and correct copy of the original document maintained as part of the activities of this office.

Form 106-I8000

  
\_\_\_\_\_  
Signature      11-8-21  
Date