



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) MIDDLETON, DANA R		
Location MORTON COUNTY / MANDAN	Intoxilyzer® Serial Number 80-005957	
Gas Lot Number 26021080A1	Gas Cylinder Number 4	Gas Expiration Date 10/05/23

Check When Done: **0 CEE 3/18/22 26021080A1**

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.080 AC
 B. 0.081 AC
 C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Dana R. Middleton</i>	Date 3/11/22
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Elder</i>	Date 3/18/22

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = MAND 8164.14.00 09/16
03/11/2022 18:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:20
02 Std. Gas	0.080	18:21
03 Room Air	0.000	18:21
04 Std. Gas	0.081	18:21
05 Room Air	0.000	18:22
06 Std. Gas	0.080	18:22
07 Room Air	0.000	18:23

Lot No = 26021080A1
Cyl No = 4
Exp Date = 10/05/2023
County = 30 Oper No. = 079706



Operator Signature
DANA MIDDLETON

Remarks:

Form 106-I8000