



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Justin William Peltz</i>		
Location <i>WILL</i>	Intoxilyzer® Serial Number <i>80-004944</i>	
Gas Lot Number <i>CYL 017</i>	Gas Cylinder Number <i>26021080A1</i>	Gas Expiration Date <i>10-05-2023</i>

Check When Done:

CEE 3.22.22

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.082* AC
 - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Justin William Peltz #156</i>	Date <i>3-17-2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Edwards</i>	Date <i>3/22/22</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WILL 8164.14.00 09/16
03/17/2022 19:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:28
02 Std. Gas	0.082	19:29
03 Room Air	0.000	19:29
04 Std. Gas	0.082	19:29
05 Room Air	0.000	19:30
06 Std. Gas	0.082	19:30
07 Room Air	0.000	19:31

Lot No = 26021080A1
Cyl No = 17
Exp Date = 10/05/2023
County = 53 Oper No. = 130755



Operator Signature
JUSTIN PELZL

Remarks: *Install
ACA Test*

Form 106-I8000