

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-004934

Agency being assigned instrument: DUMP, DISPOSED OF

Date 3/28/22

Transferred from:

Transferred to:

Name (print) CHARLES E. EDER

Name (print) DUMP

Signature Charles E. Eder

Signature Charles E. Eder

Agency O.A.G. - CRIME LAB

Agency DUMP

Reason for transfer/remarks:

- DISPOSED OF DUE TO CONTINUOUS RFI ISSUE
- DISMANTLED FOR ELECTRONIC DISPOSAL
- ND'S EEPROM CHIP DESTROYED

Or Shipped Via:




USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____



PROPERTY DISPOSAL
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 SFN 60567 (03/2022)

Multiple quantities of like items may be listed on one form and attach a sheet with tag numbers and serial numbers if available. Use a separate form for different items.

Division Crime Laboratory	Property Location Breath Lab	Date 3/28/2022
Point of Contact Charles E. Eder	Email Address ceder@nd.gov	Point of Contact Telephone Number (701) 328-6156
Description of Property (make, model, color, etc.) Intoxilyzer 8000 Evidential Breath Alcohol Instrument CMI, Inc., Owensboro, Ky		 
Tag Number (if available) AG 3407	Serial Number (if available) 80-004934	
Date Purchased (if available) NA	Original Cost (if available) NA	
Disposal Reason Continuous issue with RFI. Contacted CMI for possible solutions. Tried all suggestions. RFI issue still occurring. Instrument would have to be sent to factory for repair which would be cost prohibitive due to the age of instrument.		
Disposal Method <input checked="" type="checkbox"/> Landfill <input checked="" type="checkbox"/> Trash Can/Dumpster <input type="checkbox"/> Electronics Disposal		
Person Disposing Property Signature 	Date 3/28/22	
Send completed form to Tasha Gerding in Finance Division		
Authorized Representative Signature	Date	