

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006504 Location: ToxL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 950 psi Regulator: 967 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check MP3070
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: 22CP41184 Lot #: 22CP41184 Exp. Date: 3/20/24 3/1/24
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: 202-103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0261080A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.086</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual Inspection
replaced O-ring on breath hose.

Instrument is acceptable to be used in the field.

Janelle Pertschilker
Breath Analyst Signature

4/1/22
Date

[Signature]
Reviewed by

04/01/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 14:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 12
Standard Expiration? 11/05/2021
Oper No? 132128

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 05/31/2017
 3um 9um

	3um	9um
0th Coef(*100):	-21488	-16160
1st Coef(*100):	265581	137324
2nd Coef(*100):	2507	1149
H2O adj(mg/l*10k):	519	348

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 14:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Sol.	0.020	14:56
03 Room Air	0.000	14:57
04 Std. Sol.	0.020	14:58
05 Room Air	0.000	14:58
06 Std. Sol.	0.021	14:59
07 Room Air	0.000	14:59

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *LOW AC 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 15:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:02
02 Std. Sol.	0.401	15:04
03 Room Air	0.000	15:05
04 Std. Sol.	0.402	15:05
05 Room Air	0.000	15:06
06 Std. Sol.	0.402	15:07
07 Room Air	0.000	15:07

08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: High AC ~~0.040~~ 0.400 *for 205 July 22*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 15:09

Test	AC	Time
01 Room Air	0.000	15:10
02 *Subject Test	INT*	15:11
03 Room Air	0.000	15:11

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 15:16

Test	AC	Time
01 Diagnostic	OK	15:17
02 Room Air	0.000	15:18
03 Reference	RFI*	15:18
04 Room Air	0.000	15:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = ~~RFI~~ CHECK
Dr. Lic. = ND/DIS821456 *JVP RFI*
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 15:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:20
02 Std. Gas	0.081	15:20
03 Room Air	0.000	15:21
04 Std. Gas	0.081	15:21
05 Room Air	0.000	15:22
06 Std. Gas	0.080	15:22
07 Room Air	0.000	15:23

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller
Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry cal check 1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 15:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:24
02 Std. Gas	0.080	15:25
03 Room Air	0.000	15:25
04 Std. Gas	0.080	15:26
05 Room Air	0.000	15:26
06 Std. Gas	0.080	15:27
07 Room Air	0.000	15:27

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry cal check 4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.14.00 09/16
04/01/2022 15:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Gas	0.080	15:28
03 Room Air	0.000	15:29
04 Std. Gas	0.081	15:29
05 Room Air	0.000	15:30
06 Std. Gas	0.080	15:30
07 Room Air	0.000	15:30

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry cal check 7-9*