

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005937 Location: TOXL

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 919 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
 - 1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.021 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 01.11.24
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.309 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12.08.22

Configure dry gas standard for the following tests (Level 1,S).

- 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DL384A Lot #: 1CS6 Exp. Date: N/A
 Display reads "Interferent Detect"
- 3. RFI Check (CMS Mode) Display reads "RFI Detect"
- 4. Dry Calibration Check (Level 1,C): Known Value 0.08 AC
Gas Cyl Lot #: 05620080A1 Cyl #: 2 Exp. Date: 4/5/2022

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION
REPLACED BREATH TUBE O-RING

Instrument is acceptable to be used in the field.

[Signature]
Breath Analyst Signature

04/01/22
Date

Janelle Pertschiller
Reviewed by

04/01/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 14:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 2
Standard Expiration? 04/05/2022
Oper No? 133162

Flow Cal. Date: 11/17/2017
Slope 606
Intercept -534563

IR Calibration Date: 11/17/2017
3um 9um

0th Coef(*100): -30534 -23772
1st Coef(*100): 294336 142357
2nd Coef(*100): -906 850
H2O adj(mg/l*10k): 205 250

***** Printer Test End *****



Operator Signature

ANGELA QUINN

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 14:32


WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:33
02 Std. Sol.	0.021	14:34
03 Room Air	0.000	14:34
04 Std. Sol.	0.021	14:35
05 Room Air	0.000	14:36
06 Std. Sol.	0.021	14:36
07 Room Air	0.000	14:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08

Oper No. = 133162

 040122

Operator Signature
ANGELA QUINN

Remarks:

Cal
LOW AC Check 0.020

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 14:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Sol.	0.307	14:41
03 Room Air	0.000	14:41
04 Std. Sol.	0.309	14:42
05 Room Air	0.000	14:42
06 Std. Sol.	0.309	14:43
07 Room Air	0.000	14:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3060 *MP3064*
Std Sol No = ~~202201A~~ *202012A* *APR 04/01/22*
County = 08 Oper No. = 133162


Operator Signature *04/01/22*
ANGELA QUINN

Remarks: *HIGH AC CAL Check*

Form 106-I8000
0.030
0.300
JPL 07/21/22

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 14:52

Test	AC	Time
01 Room Air	0.000	14:53
02 *Subject Test	INT*	14:54
03 Room Air	0.000	14:54

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 05620080A1
Cyl No = 2
Expiration Date = 04/05/2022
County = 08 Oper No. = 133162



04/01/22

Operator Signature
ANGELA QUINN

Remarks:

INTERFERENT CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 14:55

Test	AC	Time
01 Diagnostic	OK	14:56
02 Room Air	RFI*	14:56
03 Room Air	0.000	14:57

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI
Dr. Lic. = ND/TES989643
Lot No = 05620080A1
Cyl No = 2
Expiration Date = 04/05/2022
County = 08 Oper No. = 133162

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
ANGELA QUINN

Remarks:

RFI

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:57
02 Std. Gas	0.079	14:58
03 Room Air	0.000	14:58
04 Std. Gas	0.079	14:59
05 Room Air	0.000	14:59
06 Std. Gas	0.080	14:59
07 Room Air	0.000	15:00

Lot No = 05620080A1
Cyl No = 2
Exp Date = 04/05/2022
County = 08

Oper No. = 133162


Operator Signature
ANGELA QUINN

04/01/22

Remarks:

DRY CAL CHECK #1-#3

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:03
02 Std. Gas	0.079	15:03
03 Room Air	0.000	15:04
04 Std. Gas	0.080	15:04
05 Room Air	0.000	15:05
06 Std. Gas	0.079	15:05
07 Room Air	0.000	15:06

Lot No = 05620080A1

Cyl No = 2

Exp Date = 04/05/2022

County = 08

Oper No. = 133162

 04/01/22

Operator Signature
ANGELA QUINN

Remarks:

Dry Cal Check #4 - #6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.14.00 09/16
04/01/2022 15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Gas	0.080	15:08
03 Room Air	0.000	15:09
04 Std. Gas	0.080	15:09
05 Room Air	0.000	15:10
06 Std. Gas	0.080	15:10
07 Room Air	0.000	15:11

Lot No = 05620080A1

Cyl No = 2

Exp Date = 04/05/2022

County = 08

Oper No. = 133162



Operator Signature
ANGELA QUINN

Remarks:

Dry Cal Check #7 - #9

Form 106-I8000