

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004940

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 959 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: ~~2021E9~~
202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: IC56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 2621080A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1	<u>0.078</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.080</u>	Test 6	<u>0.078</u>	Test 9	<u>0.080</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: Annual Inspection.

Replaced breath tube o-ring and simulator return o-ring.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

22 Apr 23
Date

[Signature]
Reviewed by

05/922
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 1
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 04/06/2020
Slope 635
Intercept -585462

IR Calibration Date: 04/06/2020
 3um 9um

0th Coef(*100): -17993 -22378
1st Coef(*100): 260246 135868
2nd Coef(*100): 2982 1272
H2O adj(mg/l*10k): 383 361

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:20
02 Std. Sol.	0.019	09:21
03 Room Air	0.000	09:22
04 Std. Sol.	0.019	09:22
05 Room Air	0.000	09:23
06 Std. Sol.	0.019	09:24
07 Room Air	0.000	09:24

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:25
02 Std. Sol.	0.402	09:26
03 Room Air	0.000	09:26
04 Std. Sol.	0.403	09:27
05 Room Air	0.000	09:28
06 Std. Sol.	0.402	09:28
07 Room Air	0.000	09:29

08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.400*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:30

Test	AC	Time
01 Room Air	0.000	09:31
02 *Subject Test	INT*	09:31
03 Room Air	0.000	09:32

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Interferent check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:32

Test	AC	Time
01 Diagnostic	OK	09:33
02 Room Air	0.000	09:34
03 *Subject Test	RFI*	09:34
04 Room Air	0.000	09:34

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:37
02 Std. Gas	0.078	09:37
03 Room Air	0.000	09:38
04 Std. Gas	0.079	09:38
05 Room Air	0.000	09:39
06 Std. Gas	0.080	09:39
07 Room Air	0.000	09:39

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:40
02 Std. Gas	0.079	09:40
03 Room Air	0.000	09:41
04 Std. Gas	0.079	09:41
05 Room Air	0.000	09:42
06 Std. Gas	0.078	09:42
07 Room Air	0.000	09:43

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.14.00 09/16
04/22/2022 09:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Gas	0.080	09:45
03 Room Air	0.000	09:46
04 Std. Gas	0.079	09:46
05 Room Air	0.000	09:47
06 Std. Gas	0.080	09:47
07 Room Air	0.000	09:48

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*