



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <i>80-004200</i>	Instrument Location <i>CARS</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <i>Annual Inspection</i>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- yes* → 6. Run Tests:
 - 01/21/22* A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *None*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Erny Dillon</i>	Date <i>04/22/2022</i>
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Pitschler</i>	Certified Date <i>06/27/2022</i>
---	-------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CARS 8164.14.00 09/16
04/22/2022 12:54

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 6
Standard Expiration? 10/05/2023
Oper No? 132318

Flow Cal. Date: 06/25/2019
Slope 614
Intercept -551153

IR Calibration Date: 06/25/2019
 3um 9um

0th Coef(*100): -18792 -22164
1st Coef(*100): 267001 140777
2nd Coef(*100): 1903 1083
H2O adj(mg/l*10k): 403 445

***** Printer Test End *****



Operator Signature
TERRY DILLON

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CARS 8164.14.00 09/16
04/22/2022 12:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:00
02 Std. Gas	0.082	13:01
03 Room Air	0.000	13:01
04 Std. Gas	0.082	13:02
05 Room Air	0.000	13:02
06 Std. Gas	0.081	13:02
07 Room Air	0.000	13:03

Lot No = 26021080A1
Cyl No = 6
Exp Date = 10/05/2023
County = 19 Oper No. = 132318



Operator Signature
TERRY DILLON

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CARS 8164.14.00 09/16
04/22/2022 13:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	13:09
02 Room Air	0.000	13:09

*Invalid Test
Inhibited - RFI

Lot No = 26021080A1
Cyl No = 6
Exp Date = 10/05/2023
County = 19 Oper No. = 132318



Operator Signature
TERRY DILLON

Remarks: