

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004939

Location: ToxL

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 523 psi Regulator: 525 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
 - 1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/2024
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 4/9/2022
 - Configure dry gas standard for the following tests (Level 1,S).
 - 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1056 Exp. Date: NA
 Display reads "Interferent Detect"
 - 3. RFI Check (CMS Mode) Display reads "RFI Detect"
 - 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1 <u>0.082</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.083</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: Annual Inspection.
Replaced breath tube o-ring + simulator return o-ring.
External Battery was replaced.

Instrument is acceptable to be used in the field.

Janelle Portscheller
Breath Analyst Signature

16 May 22
Date

[Signature]
Reviewed by

060222
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 11:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 3
Standard Expiration? 02/05/2021
Oper No? 132128

Flow Cal. Date: 01/18/2010
Slope 724
Intercept -779308

IR Calibration Date: 04/05/2017
 3um 9um

0th Coef(*100): -24406 -20805
1st Coef(*100): 261875 135290
2nd Coef(*100): 2828 1427
H2O adj(mg/l*10k): 563 393

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 11:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Sol.	0.020	11:56
03 Room Air	0.000	11:57
04 Std. Sol.	0.021	11:58
05 Room Air	0.000	11:58
06 Std. Sol.	0.021	11:59
07 Room Air	0.000	12:00

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 12:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Sol.	0.409	12:04
03 Room Air	0.000	12:05
04 Std. Sol.	0.409	12:05
05 Room Air	0.000	12:06
06 Std. Sol.	0.409	12:07
07 Room Air	0.000	12:07

08 Sim Temp = 34.0°C

Simul Ser No = MP3065
Std Sol No = 202103E
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.400*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 12:19

Test	AC	Time
01 Room Air	0.000	12:20
02 *Subject Test	INT*	12:21
03 Room Air	0.000	12:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Interferent Check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 12:22

Test	AC	Time
01 Diagnostic	OK	12:23
02 Room Air	0.000	12:23
03 *Subject Test	RFI*	12:24
04 Room Air	0.000	12:24

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: ~~DRP~~ RFI Check


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 12:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:25
02 Std. Gas	0.082	12:25
03 Room Air	0.000	12:26
04 Std. Gas	0.082	12:26
05 Room Air	0.000	12:27
06 Std. Gas	0.083	12:27
07 Room Air	0.000	12:28

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 12:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:33
02 Std. Gas	0.081	12:33
03 Room Air	0.000	12:34
04 Std. Gas	0.081	12:34
05 Room Air	0.000	12:35
06 Std. Gas	0.082	12:35
07 Room Air	0.000	12:35

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = TOXL 8164.14.00 09/16
05/16/2022 12:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Gas	0.082	12:37
03 Room Air	0.000	12:37
04 Std. Gas	0.082	12:37
05 Room Air	0.000	12:38
06 Std. Gas	0.082	12:38
07 Room Air	0.000	12:39

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*