

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004938

Location: Tox L

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 541 psi Regulator: 525 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 4/8/2022 ^{3/1/2024}
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: 262-103E Exp. Date: 4/8/2022 ^{3/24/2023}

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621086A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual Inspection.

Replaced o-ring on breath tube. Replaced o-ring on simulator return.

Instrument is acceptable to be used in the field.

Janelle Pertscheller
Breath Analyst Signature

5/16/2022
Date

[Signature]
Reviewed by

060622
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:04

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 61
Standard Expiration? 02/05/2021
Oper No? 132128

Flow Cal. Date: 06/10/2020
Slope 587
Intercept -338327

IR Calibration Date: 06/10/2020

	3um	9um
0th Coef(*100):	-22434	-23179
1st Coef(*100):	269849	136790
2nd Coef(*100):	963	840
H2O adj(mg/l*10k):	355	370

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:06
02 Std. Sol.	0.019	01:08
03 Room Air	0.000	01:08
04 Std. Sol.	0.019	01:09
05 Room Air	0.000	01:09
06 Std. Sol.	0.019	01:10
07 Room Air	0.000	01:11

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:14
02 Std. Sol.	0.389	01:15
03 Room Air	0.000	01:16
04 Std. Sol.	0.392	01:16
05 Room Air	0.000	01:17
06 Std. Sol.	0.393	01:18
07 Room Air	0.000	01:18

08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.400*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:22

Test	AC	Time
01 Room Air	0.000	01:23
02 *Subject Test	INT*	01:23
03 Room Air	0.000	01:24

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:24

Test	AC	Time
01 Diagnostic	OK	01:25
02 Room Air	0.000	01:25
03 *Subject Test	RFI*	01:26
04 Room Air	0.000	01:26

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:27
02 Std. Gas	0.078	01:27
03 Room Air	0.000	01:28
04 Std. Gas	0.079	01:28
05 Room Air	0.000	01:29
06 Std. Gas	0.079	01:29
07 Room Air	0.000	01:30

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 01:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	02:00
02 Std. Gas	0.079	02:00
03 Room Air	0.000	02:01
04 Std. Gas	0.079	02:01
05 Room Air	0.000	02:02
06 Std. Gas	0.079	02:02
07 Room Air	0.000	02:03

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check # 4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
05/16/2022 02:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	02:31
02 Std. Gas	0.079	02:31
03 Room Air	0.000	02:31
04 Std. Gas	0.079	02:32
05 Room Air	0.000	02:32
06 Std. Gas	0.079	02:33
07 Room Air	0.000	02:33

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*