

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004956

Location: TDXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 960 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.20 AC
Sim. Ser #: MP3060 Lot #: 2022.01A Exp. Date: 1/11/2024
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 2020 12A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0262108DA1 Cyl #: 4 Exp. Date: 3/5/2023

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual Inspection
Replaced breath tube o-ring. Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Putschillor
Breath Analyst Signature

5/16/2022
Date

[Signature]
Reviewed by

061522
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.14.00 09/16
05/16/2022 12:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:28
02 Std. Sol.	0.301	12:34
03 Room Air	0.000	12:34
04 Std. Sol.	0.303	12:35
05 Room Air	0.000	12:36
06 Std. Sol.	0.303	12:36
07 Room Air	0.000	12:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.14.00 09/16
05/16/2022 12:41

Test	AC	Time
01 Diagnostic	OK	12:43
02 Room Air	0.000	12:43
03 *Subject Test	RFI*	12:44
04 Room Air	0.000	12:45

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.14.00 09/16
05/16/2022 12:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:51
02 Std. Gas	0.081	12:51
03 Room Air	0.000	12:52
04 Std. Gas	0.081	12:52
05 Room Air	0.000	12:53
06 Std. Gas	0.081	12:53
07 Room Air	0.000	12:54


Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.14.00 09/16
05/16/2022 12:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Gas	0.081	13:00
03 Room Air	0.000	13:00
04 Std. Gas	0.081	13:00
05 Room Air	0.000	13:01
06 Std. Gas	0.081	13:01
07 Room Air	0.000	13:02

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.14.00 09/16
05/16/2022 13:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:04
02 Std. Gas	0.082	13:05
03 Room Air	0.000	13:05
04 Std. Gas	0.081	13:06
05 Room Air	0.000	13:06
06 Std. Gas	0.081	13:07
07 Room Air	0.000	13:07

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*