

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004948

Location: ToxL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 494 psi Regulator: 525 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/2024
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/2022

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3047 Lot #: K56 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 4 Exp. Date: 3/5/2023

Test 1 <u>0.083</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: Replaced o-ring on breath tube.  
Annual Inspection.

Instrument is acceptable to be used in the field.

Janille Postschaller  
Breath Analyst Signature

5/16/2022  
Date

[Signature]  
Reviewed by

060622  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004948  
Location = TOXL                    8164.14.00 09/16  
05/16/2022                            11:15

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=-|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    34418080A2  
Standard Cyl #?                    52  
Standard Expiration?                02/05/2021  
Oper No?                            132128

Flow Cal. Date:                    08/28/2018  
Slope                                642  
Intercept                            -606282

IR Calibration Date:                08/28/2018  
    3um                                9um

-----  
0th Coef(\*100):            -13609                            -14548  
1st Coef(\*100):            248186                            128726  
2nd Coef(\*100):            3538                                1508  
H2O adj (mg/l\*10k):        602                                515

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.14.00 09/16  
05/16/2022      11:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Sol.	0.021	11:21
03 Room Air	0.000	11:22
04 Std. Sol.	0.021	11:23
05 Room Air	0.000	11:23
06 Std. Sol.	0.021	11:24
07 Room Air	0.000	11:24

08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 202201A

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC ~~0.2000~~ 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.14.00 09/16  
05/16/2022      11:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:27
02 Std. Sol.	0.296	11:28
03 Room Air	0.000	11:28
04 Std. Sol.	0.297	11:29
05 Room Air	0.000	11:30
06 Std. Sol.	0.298	11:30
07 Room Air	0.000	11:31

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC 0.300*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004948  
Location = TOXL                 8164.14.00 09/16  
05/16/2022                                 11:48

Test	AC	Time
01 Room Air	0.000	11:49
02 *Subject Test	INT*	11:50
03 Room Air	0.000	11:50

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                                 Weight = 150  
Test = DUI                                 Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 4  
Expiration Date = 03/05/2023  
County = 08                                 Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004948  
Location = TOXL                 8164.14.00 09/16  
05/16/2022                        11:52

Test	AC	Time
01 Diagnostic	OK	11:53
02 Room Air	0.000	11:53
03 *Subject Test	RFI*	11:55
04 Room Air	0.000	11:55

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 4  
Expiration Date = 03/05/2023  
County = 08                          Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004948  
Location = TOXL                    8164.14.00 09/16  
05/16/2022                            11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Gas	0.083	11:56
03 Room Air	0.000	11:57
04 Std. Gas	0.082	11:57
05 Room Air	0.000	11:57
06 Std. Gas	0.082	11:58
07 Room Air	0.000	11:58


Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Checks #1-3*

Form 106-I8000





Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004948  
Location = TOXL                    8164.14.00 09/16  
05/16/2022                            12:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Gas	0.082	12:04
03 Room Air	0.000	12:04
04 Std. Gas	0.082	12:04
05 Room Air	0.000	12:05
06 Std. Gas	0.082	12:05
07 Room Air	0.000	12:06

Lot No = 02621080A1  
Cyl No = 4  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*