

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004954

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 517 psi Regulator: 525 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP6636 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3069 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1C56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621085 Cyl #: 2 Exp. Date: 3/5/23

| | | | | | |
|---------|--------------|--------|--------------|--------|--------------|
| Test 1 | <u>0.083</u> | Test 4 | <u>0.083</u> | Test 7 | <u>0.082</u> |
| Test 2 | <u>0.082</u> | Test 5 | <u>0.082</u> | Test 8 | <u>0.083</u> |
| Test 3 | <u>0.082</u> | Test 6 | <u>0.083</u> | Test 9 | <u>0.082</u> |
| Average | <u>0.082</u> | | | | |

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring and simulator return o-ring.

Instrument is acceptable to be used in the field.

Janelle Patschiller
Breath Analyst Signature

5/17/22
Date

Reviewed by

060322
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 08:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 2
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 08/27/2012
Slope 670
Intercept -570610

IR Calibration Date: 05/31/2017
 3um 9um

0th Coef(*100): -18765 -20238
1st Coef(*100): 256075 130805
2nd Coef(*100): 2112 1219
H2O adj(mg/l*10k): 442 407

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 09:00

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 09:01 |
| 02 Std. Sol. | 0.018 | 09:02 |
| 03 Room Air | 0.000 | 09:03 |
| 04 Std. Sol. | 0.019 | 09:04 |
| 05 Room Air | 0.000 | 09:04 |
| 06 Std. Sol. | 0.020 | 09:05 |
| 07 Room Air | 0.000 | 09:05 |


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 09:24

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 09:25 |
| 02 Std. Sol. | 0.298 | 09:26 |
| 03 Room Air | 0.000 | 09:27 |
| 04 Std. Sol. | 0.298 | 09:27 |
| 05 Room Air | 0.000 | 09:28 |
| 06 Std. Sol. | 0.298 | 09:29 |
| 07 Room Air | 0.000 | 09:29 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 202201F
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac Check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 09:37

| Test | AC | Time |
|------------------|-------|-------|
| 01 Diagnostic | OK | 09:37 |
| 02 Room Air | 0.000 | 09:38 |
| 03 *Subject Test | RFI* | 09:38 |
| 04 Room Air | 0.000 | 09:39 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: ~~Interferent Check~~
RFI check


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 11:49

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:50 |
| 02 Std. Gas | 0.083 | 11:50 |
| 03 Room Air | 0.000 | 11:50 |
| 04 Std. Gas | 0.082 | 11:51 |
| 05 Room Air | 0.000 | 11:51 |
| 06 Std. Gas | 0.082 | 11:52 |
| 07 Room Air | 0.000 | 11:52 |

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 11:53

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:54 |
| 02 Std. Gas | 0.083 | 11:54 |
| 03 Room Air | 0.000 | 11:55 |
| 04 Std. Gas | 0.082 | 11:55 |
| 05 Room Air | 0.000 | 11:55 |
| 06 Std. Gas | 0.083 | 11:56 |
| 07 Room Air | 0.000 | 11:56 |

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
05/17/2022 11:57

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:57 |
| 02 Std. Gas | 0.082 | 11:58 |
| 03 Room Air | 0.000 | 11:58 |
| 04 Std. Gas | 0.083 | 11:58 |
| 05 Room Air | 0.000 | 11:59 |
| 06 Std. Gas | 0.082 | 11:59 |
| 07 Room Air | 0.000 | 12:00 |

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*