

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004935

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 966 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/2024
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/2023

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC

20 July 22
Gas Cyl Lot #: 02621080 A1 Cyl #: 31 Exp. Date: 3/5/2023
Test 1 0.077 Test 4 0.077 Test 7 0.078
Test 2 0.078 Test 5 0.077 Test 8 0.078
Test 3 0.078 Test 6 0.077 Test 9 0.077
Average 0.077

C. Remarks/Maintenance: Annual Inspection.

Replaced breath tube o-ring and simulator return o-ring.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Putschillo
Breath Analyst Signature

5/17/2022
Date

[Signature]
Reviewed by

060622
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 11:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 9
Standard Expiration? 03/05/2019
Oper No? 132128

Flow Cal. Date: 07/12/2012
Slope 655
Intercept -510289

IR Calibration Date: 05/20/2015
 3um 9um

0th Coef(*100): -7871 -19684
1st Coef(*100): 264505 133108
2nd Coef(*100): 2865 1320
H2O adj(mg/l*10k): 200 298

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 11:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Sol.	0.018	11:48
03 Room Air	0.000	11:49
04 Std. Sol.	0.017	11:49
05 Room Air	0.000	11:50
06 Std. Sol.	0.018	11:50
07 Room Air	0.000	11:51

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 11:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:54
02 Std. Sol.	0.397	11:56
03 Room Air	0.000	11:56
04 Std. Sol.	0.399	11:57
05 Room Air	0.000	11:57
06 Std. Sol.	0.399	11:58
07 Room Air	0.000	11:59

08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

High AC check 0.400

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 13:08

Test	AC	Time
01 Diagnostic	OK	13:09
02 Room Air	0.000	13:10
03 *Subject Test	RFI*	13:10
04 Room Air	0.000	13:10

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.077	13:12
03 Room Air	0.000	13:12
04 Std. Gas	0.078	13:12
05 Room Air	0.000	13:13
06 Std. Gas	0.078	13:13
07 Room Air	0.000	13:14

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 13:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Gas	0.077	13:24
03 Room Air	0.000	13:25
04 Std. Gas	0.077	13:25
05 Room Air	0.000	13:26
06 Std. Gas	0.077	13:26
07 Room Air	0.000	13:27

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.14.00 09/16
05/17/2022 13:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Gas	0.078	13:28
03 Room Air	0.000	13:28
04 Std. Gas	0.078	13:28
05 Room Air	0.000	13:29
06 Std. Gas	0.077	13:29
07 Room Air	0.000	13:30

Lot No = 02621080A1

Cyl No = 31

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 7-9*