

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: Intox 8000 Serial Number: 80-005947

Agency being assigned instrument: ND0AG - CLD 20 July 22
~~Dickinson Police Department~~

Date 5/17/22

Transferred from:

Name (print) Kevin Eldridge

Signature [Signature]

Agency STARK CO SO

Transferred to:

Name (print) Janelle Pertscheller

Signature [Signature]

Agency ND0AG - CLD

Reason for transfer/remarks:

for 5/17/22
Repair + annual inspection. EPROM error
message on instrument. Replace with
80-004939

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____