



ETHANOL GAS STANDARD CYLINDER REPORT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Czapiewski, Rick D</i>		
Location <i>FINL</i>	Intoxilyzer® Serial Number <i>80-005946</i>	
Gas Lot Number <i>26021080A1</i>	Gas Cylinder Number <i>22</i>	Gas Expiration Date <i>10/5/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. O. *82* AC
 - B. O. *80* AC
 - C. O. *81* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Rick Czapiewski</i>	Date <i>5/20/2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Putschillor</i>	Date <i>27 May 22</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = FINL 8164.14.00 09/16
05/20/2022 11:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Gas	0.082	11:36
03 Room Air	0.000	11:37
04 Std. Gas	0.080	11:37
05 Room Air	0.000	11:38
06 Std. Gas	0.081	11:38
07 Room Air	0.000	11:39

Lot No = 26021080A1

Cyl No = 22

Exp Date = 10/05/2023

County = 46

Oper No. = 130363



Operator Signature
RICK CZAPIEWSKI

Remarks:

Form 106-I8000