



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Matthew E (M)</i>		
Location <i>WCPD</i>	Intoxilyzer® Serial Number <i>SN 80-004956</i>	
Gas Lot Number <i>333 21080AH</i>	Gas Cylinder Number <i>011</i>	Gas Expiration Date <i>1-5-2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.082*-AC
 B. *0.082*-AC
 C. *0.082*-AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Matthew E. Carl</i>	Date <i>05/20/22</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Putschler</i>	Date <i>06 June 22</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = WCPD 8164.14.00 09/16
05/20/2022 11:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:32
02 Std. Gas	0.082	11:32
03 Room Air	0.000	11:33
04 Std. Gas	0.082	11:33
05 Room Air	0.000	11:34
06 Std. Gas	0.082	11:34
07 Room Air	0.000	11:34

Lot No = 33321080A4

Cyl No = 11

Exp Date = 01/05/2024

County = 27

Oper No. = 133050



Operator Signature
MATTHEW EARL

Remarks: *INST 11/17*