



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-005946	Instrument Location FINL
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Install after being at lab for maint.</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 5/20/2022
-------------------------------	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration for the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 27 May 22
---------------------------	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = FINL 8164.14.00 09/16
05/20/2022 11:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 22
Standard Expiration? 10/05/2023
Oper No? 130363

Flow Cal. Date: 06/02/2021
Slope 642
Intercept -600927

IR Calibration Date: 06/02/2021
3um 9um

	3um	9um
0th Coef(*100):	-20165	-16483
1st Coef(*100):	290371	141330
2nd Coef(*100):	953	910
H2O adj(mg/l*10k):	496	404

***** Printer Test End *****



Operator Signature
RICK CZAPIEWSKI

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = FINL 8164.14.00 09/16
05/20/2022 11:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Gas	0.080	11:47
03 Room Air	0.000	11:47
04 Std. Gas	0.081	11:48
05 Room Air	0.000	11:48
06 Std. Gas	0.081	11:49
07 Room Air	0.000	11:49

Lot No = 26021080A1

Cyl No = 22

Exp Date = 10/05/2023

County = 46

Oper No. = 130363



Operator Signature
RICK CZAPIEWSKI

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = FINL 8164.14.00 09/16
05/20/2022 11:51

Test	AC	Time
01 Room Air	RFI*	11:55
02 Room Air	RFI*	11:55

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, NA NA
Sub DOB = 11/11/2000
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 26021080A1
Cyl No = 22
Expiration Date = 10/05/2023
County = 46 Oper No. = 130363



Operator Signature
RICK CZAPIEWSKI

Remarks:

RFI