

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 86-004184

Location: ToxL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 973 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3086 Lot #: 202201F Exp. Date: 01/18/24
306B

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5440 Lot #: 1001156 Exp. Date: NA
DR3847
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0262.1080A1 Cyl #: 2 Exp. Date: 3/5/23
Test 1 0.080 Test 4 0.080 Test 7 0.081
Test 2 0.080 Test 5 0.080 Test 8 0.080
Test 3 0.080 Test 6 0.080 Test 9 0.081
Average 0.080

C. Remarks/Maintenance: Annual Inspection

Replaced breath tube o-ring and simulator return o-ring.

Instrument is acceptable to be used in the field.

Janelle Portschiller
Breath Analyst Signature

02 June 22
Date

[Signature]
Reviewed by

060222
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 24
Standard Expiration? 10/05/2023
Oper No? 132128

Flow Cal. Date: 07/30/2010
Slope 804
Intercept -925008

IR Calibration Date: 12/07/2015
 3um 9um

0th Coef(*100): -15679 -18680
1st Coef(*100): 263558 136948
2nd Coef(*100): 2594 1146
H2O adj(mg/l*10k): 405 327

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:19
02 Std. Sol.	0.018	09:19
03 Room Air	0.000	09:20
04 Std. Sol.	0.019	09:21
05 Room Air	0.000	09:21
06 Std. Sol.	0.018	09:22
07 Room Air	0.000	09:22

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 21340
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:25
02 Std. Sol.	0.294	09:26
03 Room Air	0.000	09:27
04 Std. Sol.	0.296	09:28
05 Room Air	0.000	09:28
06 Std. Sol.	0.297	09:29
07 Room Air	0.000	09:30

08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 202201F

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Hg in AC 0.300


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:39

Test	AC	Time
01 Room Air	0.000	09:40
02 *Subject Test	INT*	09:41
03 Room Air	0.000	09:41

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:42

Test	AC	Time
01 Diagnostic	OK	09:43
02 Room Air	0.000	09:43
03 *Subject Test	RFI*	09:43
04 Room Air	0.000	09:44

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Gas	0.080	09:45
03 Room Air	0.000	09:46
04 Std. Gas	0.080	09:46
05 Room Air	0.000	09:47
06 Std. Gas	0.080	09:47
07 Room Air	0.000	09:47

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:48
02 Std. Gas	0.080	09:49
03 Room Air	0.000	09:49
04 Std. Gas	0.080	09:50
05 Room Air	0.000	09:50
06 Std. Gas	0.080	09:51
07 Room Air	0.000	09:51

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.14.00 09/16
06/02/2022 09:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:53
02 Std. Gas	0.081	09:54
03 Room Air	0.000	09:54
04 Std. Gas	0.080	09:54
05 Room Air	0.000	09:55
06 Std. Gas	0.081	09:55
07 Room Air	0.000	09:56

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*