

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006491 Location: TUXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1001 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0262108A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual inspection
Replaced breath tube o-ring.

Instrument is acceptable to be used in the field.

Janelle Putschiller
Breath Analyst Signature

6/8/22
Date

[Signature]
Reviewed by

060922
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:22
02 Std. Sol.	0.017	15:23
03 Room Air	0.000	15:23
04 Std. Sol.	0.018	15:24
05 Room Air	0.000	15:25
06 Std. Sol.	0.018	15:25
07 Room Air	0.000	15:26

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 21340
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Sol.	0.303	15:28
03 Room Air	0.000	15:29
04 Std. Sol.	0.303	15:30
05 Room Air	0.000	15:31
06 Std. Sol.	0.302	15:31
07 Room Air	0.000	15:32


08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 202201F

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:33

Test	AC	Time
01 Room Air	0.000	15:34
02 *Subject Test	INT*	15:35
03 Room Air	0.000	15:35

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:36

Test	AC	Time
01 Diagnostic	OK	15:36
02 Room Air	0.000	15:37
03 *Subject Test	RFI*	15:37
04 Room Air	0.000	15:38

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:39
02 Std. Gas	0.081	15:39
03 Room Air	0.000	15:40
04 Std. Gas	0.081	15:40
05 Room Air	0.000	15:41
06 Std. Gas	0.079	15:41
07 Room Air	0.000	15:42

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:43
02 Std. Gas	0.080	15:43
03 Room Air	0.000	15:44
04 Std. Gas	0.081	15:44
05 Room Air	0.000	15:44
06 Std. Gas	0.081	15:45
07 Room Air	0.000	15:45

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = TOXL 8164.14.00 09/16
06/08/2022 15:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:46
02 Std. Gas	0.081	15:47
03 Room Air	0.000	15:47
04 Std. Gas	0.080	15:48
05 Room Air	0.000	15:48
06 Std. Gas	0.081	15:48
07 Room Air	0.000	15:49

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low 9th Dry Calibration Check #7-9*