

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005951 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 941 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
    - b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.300 AC  
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 020 Exp. Date: 3/5/23

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection.  
Replaced breath tube O-ring. Replaced elbow connection  
for calibration inlet. +HPV

Instrument is acceptable to be used in the field.

Janelle Portschiller  
Breath Analyst Signature

6/9/2022  
Date

[Signature]  
Reviewed by

06/10/22  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-005951  
Location = TOXL                   8164.14.00 09/16  
06/09/2022                                 15:06

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?            No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?            0.080  
Standard Lot #?            26021080A1  
Standard Cyl #?            29  
Standard Expiration?       10/05/2023  
Oper No?                    132128

Flow Cal. Date:            06/21/2013  
Slope                       693  
Intercept                  -721277

IR Calibration Date:       05/08/2017  
                              3um                    9um

	3um	9um
0th Coef(*100):	-17201	-17496
1st Coef(*100):	267144	137813
2nd Coef(*100):	2948	1389
H2O adj(mg/l*10k):	438	428

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005951  
Location = TOXL      8164.14.00 09/16  
06/09/2022      15:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:16
02 Std. Sol.	0.019	15:17
03 Room Air	0.000	15:18
04 Std. Sol.	0.020	15:19
05 Room Air	0.000	15:19
06 Std. Sol.	0.020	15:20
07 Room Air	0.000	15:20

08 Sim Temp = 34.0°C

Simul Ser No = MP3060  
Std Sol No = 202201A  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005951  
Location = TOXL                    8164.14.00 09/16  
06/09/2022                            15:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:24
02 Std. Sol.	0.303	15:25
03 Room Air	0.000	15:25
04 Std. Sol.	0.305	15:26
05 Room Air	0.000	15:27
06 Std. Sol.	0.308	15:27
07 Room Air	0.000	15:28


08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: <sup>High</sup>  
~~Low~~ AC Check 0.300

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005951  
Location = TOXL                8164.14.00 09/16  
06/09/2022                        15:29

Test	AC	Time
01 Room Air	0.000	15:30
02 *Subject Test	INT*	15:30
03 Room Air	0.000	15:31

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 20  
Expiration Date = 03/05/2023  
County = 08                        Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent check*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005951  
Location = TOXL                    8164.14.00 09/16  
06/09/2022                            15:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:34
02 Std. Gas	0.081	15:34
03 Room Air	0.000	15:35
04 Std. Gas	0.081	15:35
05 Room Air	0.000	15:36
06 Std. Gas	0.080	15:36
07 Room Air	0.000	15:37

Lot No = 02621080A1  
Cyl No = 20  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *dry calibration check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005951  
Location = TOXL                      8164.14.00 09/16  
06/09/2022                              15:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:38
02 Std. Gas	0.081	15:38
03 Room Air	0.000	15:39
04 Std. Gas	0.081	15:39
05 Room Air	0.000	15:40
06 Std. Gas	0.080	15:40
07 Room Air	0.000	15:41

Lot No = 02621080A1  
Cyl No = 20  
Exp Date = 03/05/2023  
County = 08                              Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer          Alcohol Analyzer  
North Dakota Model 8000          SN 80-005951  
Location = TOXL                  8164.14.00 09/16  
06/09/2022                                  15:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:44
02 Std. Gas	0.081	15:45
03 Room Air	0.000	15:45
04 Std. Gas	0.081	15:46
05 Room Air	0.000	15:46
06 Std. Gas	0.081	15:47
07 Room Air	0.000	15:47

Lot No = 02621080A1

Cyl No = 20

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Looks Dry Calibration Check # 7-9*