

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005957

Location: TUCL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 947 psi Regulator: 975 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1C56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 20 Exp. Date: 3/5/23

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.078</u>
Test 2	<u>0.078</u>	Test 5	<u>0.079</u>	Test 8	<u>0.078</u>
Test 3	<u>0.078</u>	Test 6	<u>0.078</u>	Test 9	<u>0.079</u>
Average	<u>0.078</u> ^{20 JUL 2022}				

C. Remarks/Maintenance: Annual inspection

Replaced breath tube o-ring.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

6/9/2022
Date

[Signature]
Reviewed by

06/07/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = ~~MAND~~ *Jup* 8164.14.00 09/16
06/09/2022 *Tox L* 13:05

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 4
Standard Expiration? 10/05/2023
Oper No? 132128

Flow Cal. Date: 06/21/2013
Slope 647
Intercept -613317

IR Calibration Date: 05/11/2015

	3um	9um
0th Coef(*100):	-16666	-18045
1st Coef(*100):	278478	132949
2nd Coef(*100):	3303	1397
H2O adj(mg/l*10k):	379	351

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:25
02 Std. Sol.	0.019	13:26
03 Room Air	0.000	13:26
04 Std. Sol.	0.019	13:27
05 Room Air	0.000	13:27
06 Std. Sol.	0.018	13:28
07 Room Air	0.000	13:29

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:31
02 Std. Sol.	0.297	13:31
03 Room Air	0.000	13:32
04 Std. Sol.	0.301	13:33
05 Room Air	0.000	13:33
06 Std. Sol.	0.302	13:34
07 Room Air	0.000	13:35

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 202012A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:35

Test	AC	Time
01 Room Air	0.000	13:36
02 *Subject Test	INT*	13:37
03 Room Air	0.000	13:37

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 20
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:38

Test	AC	Time
01 Diagnostic	OK	13:38
02 Room Air	0.000	13:39
03 *Subject Test	RFI*	13:39
04 Room Air	0.000	13:39

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 20
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Gas	0.079	13:41
03 Room Air	0.000	13:41
04 Std. Gas	0.078	13:42
05 Room Air	0.000	13:42
06 Std. Gas	0.078	13:42
07 Room Air	0.000	13:43

Lot No = 02621080A1
Cyl No = 20
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Gas	0.079	13:45
03 Room Air	0.000	13:46
04 Std. Gas	0.079	13:46
05 Room Air	0.000	13:46
06 Std. Gas	0.078	13:47
07 Room Air	0.000	13:47

Lot No = 02621080A1
Cyl No = 20
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 46*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
06/09/2022 13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Gas	0.078	13:48
03 Room Air	0.000	13:49
04 Std. Gas	0.078	13:49
05 Room Air	0.000	13:50
06 Std. Gas	0.079	13:50
07 Room Air	0.000	13:51

Lot No = 02621080A1
Cyl No = 20
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*