

CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: Intoxilyzer 8000 Serial Number: 80-006506

Agency being assigned instrument: NDOAG-CLD

Date 6/13/22

Transferred from:

Transferred to:

Name (print) Terry Sand

Name (print) Deanna Dailey

Signature Terry Sand

Signature Deanna Dailey

Agency Bottineau Co. S.O.

Agency NDOAG-~~Adm~~ CLD

Reason for transfer/remarks:

To NDOAG-CLD annual inspection.  
Calibration check of dry gas were low. Will replace  
with 80-004936

Or Shipped Via:

USPS \_\_\_\_\_ UPS \_\_\_\_\_ FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_