

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006498

Location: TXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 946 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6636 Lot #: 21340 Exp. Date: 8/24/23
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.360 AC
Sim. Ser #: MP368 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: 023847 Lot #: 1056 Exp. Date: N/A
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621020A1 Cyl #: 002 Exp. Date: 3/6/23

Test 1 <u>0.079</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.079</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced electrical tape on breath tube. Replaced breath tube o-ring. Replaced external battery.

Instrument is acceptable to be used in the field.

Jennifer Portschiller
Breath Analyst Signature

6/14/22
Date

[Signature]
Reviewed by

06/15/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 26
Standard Expiration? 10/05/2023
Oper No? 132128

Flow Cal. Date: 08/18/2015
Slope 669
Intercept -637469

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100): -28161 -22508
1st Coef(*100): 286657 143843
2nd Coef(*100): 2341 1281
H2O adj(mg/l*10k): 776 451

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:22
02 Std. Sol.	0.019	10:23
03 Room Air	0.000	10:24
04 Std. Sol.	0.020	10:25
05 Room Air	0.000	10:25
06 Std. Sol.	0.020	10:26
07 Room Air	0.000	10:26

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 21340
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Sol.	0.303	10:29
03 Room Air	0.000	10:29
04 Std. Sol.	0.305	10:30
05 Room Air	0.000	10:31
06 Std. Sol.	0.305	10:31
07 Room Air	0.000	10:32

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 202201F
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.300*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:34

Test	AC	Time
01 Room Air	0.000	10:35
02 *Subject Test	INT*	10:35
03 Room Air	0.000	10:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: ~~RFI check~~ ^{Jan} Interferent check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:36

Test	AC	Time
01 Diagnostic	OK	10:37
02 Room Air	0.000	10:37
03 *Subject Test	RFI*	10:38
04 Room Air	0.000	10:39

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:40
02 Std. Gas	0.079	10:41
03 Room Air	0.000	10:41
04 Std. Gas	0.079	10:41
05 Room Air	0.000	10:42
06 Std. Gas	0.080	10:42
07 Room Air	0.000	10:43

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Checks #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:44
02 Std. Gas	0.080	10:44
03 Room Air	0.000	10:45
04 Std. Gas	0.080	10:45
05 Room Air	0.000	10:46
06 Std. Gas	0.080	10:46
07 Room Air	0.000	10:47

Lot No = 02621080A1

Cyl No = 2

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: Dry Calibration Check #1-3 ~~4-6~~

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.14.00 09/16
06/14/2022 10:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:48
02 Std. Gas	0.080	10:48
03 Room Air	0.000	10:49
04 Std. Gas	0.080	10:49
05 Room Air	0.000	10:50
06 Std. Gas	0.079	10:50
07 Room Air	0.000	10:51


Lot No = 02621080A1

Cyl No = 2

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: Dry Calibration Check # 7-9