

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006496

Location: TUCL

A. General Setup and Checks:

1. ☒ Diagnostic Tests Pass and Instrument in Ready Mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 213 psi Regulator: 228 psi

B. Tests (Sign and attach test records):

☒ Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. ☒ Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP 3067 Lot #: 21420 Exp. Date: 10/20/23 ^{MP 5290 21340}
- b. ☒ High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP 3067 Lot #: 21420 Exp. Date: 10/20/23 ^{8/24/23}

☒ Configure dry gas standard for the following tests (Level 1,S).

2. ☒ Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1C5C Exp. Date: NA

☒ Display reads "Interferent Detect"

3. ☒ RFI Check (CMS Mode)

☒ Display reads "RFI Detect"

4. ☒ Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 0262108 A1 Cyl #: 4 Exp. Date: 3/5/2023 ²⁶²³

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring + simulator return o-ring.

☒ Instrument is acceptable to be used in the field.

Janette Pertschiller
Breath Analyst Signature

6/14/22
Date

[Signature]
Reviewed by

06/15/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:00

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 20
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 01/29/2020
Slope 663
Intercept -706121

IR Calibration Date: 08/18/2015
 3um 9um

	3um	9um
0th Coef(*100):	-22081	-15963
1st Coef(*100):	278490	136818
2nd Coef(*100):	2086	1282
H2O adj(mg/l*10k):	598	386

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:03
02 Std. Sol.	0.020	20:05
03 Room Air	0.000	20:05
04 Std. Sol.	0.020	20:06
05 Room Air	0.000	20:06
06 Std. Sol.	0.020	20:07
07 Room Air	0.000	20:08

08 Sim Temp = 34.0°C

Simul Ser No = MP5290-*der*
Std Sol No = 21340
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks:

Low AC Check 0.020

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:12
02 Std. Sol.	0.299	20:13
03 Room Air	0.000	20:13
04 Std. Sol.	0.301	20:14
05 Room Air	0.000	20:15
06 Std. Sol.	0.301	20:15
07 Room Air	0.000	20:16

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:39

Test	AC	Time
01 Room Air	0.000	20:40
02 *Subject Test	INT*	20:42
03 Room Air	0.000	20:43

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:45

Test	AC	Time
01 Diagnostic	OK	20:46
02 Room Air	0.000	20:47
03 *Subject Test	RFI*	20:47
04 Room Air	0.000	20:47

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:49
02 Std. Gas	0.081	20:50
03 Room Air	0.000	20:50
04 Std. Gas	0.080	20:50
05 Room Air	0.000	20:51
06 Std. Gas	0.081	20:51
07 Room Air	0.000	20:52


Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:53
02 Std. Gas	0.081	20:53
03 Room Air	0.000	20:54
04 Std. Gas	0.081	20:54
05 Room Air	0.000	20:55
06 Std. Gas	0.082	20:55
07 Room Air	0.000	20:56

Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *dry calibration check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = TOXL 8164.14.00 09/16
06/14/2022 20:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:57
02 Std. Gas	0.081	20:57
03 Room Air	0.000	20:58
04 Std. Gas	0.082	20:58
05 Room Air	0.000	20:59
06 Std. Gas	0.081	20:59
07 Room Air	0.000	21:00

Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 7-9*

Form 106-I8000