

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006505 Location: ToxL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1001 psi Regulator: 1600 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 31202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 2 Exp. Date: 3/5/23

Test 1 <u>0.080</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection.
Replaced breath tube o-ring and simulator return o-ring.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Parkschiller
Breath Analyst Signature

6/14/22
Date

[Signature]
Reviewed by

062322
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:05
02 Std. Sol.	0.020	09:05
03 Room Air	0.000	09:06
04 Std. Sol.	0.019	09:07
05 Room Air	0.000	09:07
06 Std. Sol.	0.019	09:08
07 Room Air	0.000	09:08

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:11
02 Std. Sol.	0.302	09:12
03 Room Air	0.000	09:13
04 Std. Sol.	0.304	09:14
05 Room Air	0.000	09:14
06 Std. Sol.	0.303	09:15
07 Room Air	0.000	09:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 202201F
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:17

Test	AC	Time
01 Room Air	0.000	09:18
02 *Subject Test	INT*	09:20
03 Room Air	0.000	09:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check for
Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:21

Test	AC	Time
01 Diagnostic	OK	09:22
02 Room Air	0.000	09:23
03 *Subject Test	RFI*	09:23
04 Room Air	0.000	09:24

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Gas	0.080	09:28
03 Room Air	0.000	09:28
04 Std. Gas	0.079	09:29
05 Room Air	0.000	09:29
06 Std. Gas	0.080	09:30
07 Room Air	0.000	09:30

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:33
02 Std. Gas	0.079	09:34
03 Room Air	0.000	09:34
04 Std. Gas	0.080	09:34
05 Room Air	0.000	09:35
06 Std. Gas	0.079	09:35
07 Room Air	0.000	09:36

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006505
Location = TOXL 8164.14.00 09/16
06/14/2022 09:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:37
02 Std. Gas	0.079	09:38
03 Room Air	0.000	09:38
04 Std. Gas	0.080	09:39
05 Room Air	0.000	09:39
06 Std. Gas	0.079	09:39
07 Room Air	0.000	09:40

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*