

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006501

Location: TUOL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 989 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5296 Lot #: 21346 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).


2. Interferent Check (Level 1,B): Known Value: 0.16 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 12677CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect" 0.080
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 4 Exp. Date: 3/5/23
Test 1 0.076 Test 4 0.078 Test 7 0.078
Test 2 0.078 Test 5 0.078 Test 8 0.079
Test 3 0.077 Test 6 0.078 Test 9 0.079
Average 0.078

C. Remarks/Maintenance: Annual inspection.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

6/14/22
Date


Reviewed by

06/16/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 14:28

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 18
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 06/13/2019
Slope 651
Intercept -777821

IR Calibration Date: 06/13/2019

	3um	9um
0th Coef(*100):	-20531	-14906
1st Coef(*100):	270252	139749
2nd Coef(*100):	1917	1284
H2O adj(mg/l*10k):	393	321

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 14:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:30
02 Std. Sol.	0.020	14:31
03 Room Air	0.000	14:31
04 Std. Sol.	0.020	14:32
05 Room Air	0.000	14:33
06 Std. Sol.	0.020	14:33
07 Room Air	0.000	14:34

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 14:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Sol.	0.304	14:50
03 Room Air	0.000	14:50
04 Std. Sol.	0.305	14:51
05 Room Air	0.000	14:52
06 Std. Sol.	0.305	14:52
07 Room Air	0.000	14:53

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac Check 0.300*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 14:55

Test	AC	Time
01 Room Air	0.000	14:56
02 *Subject Test	INT*	14:58
03 Room Air	0.000	14:59

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 15:00

Test	AC	Time
01 Diagnostic	OK	15:02
02 Room Air	0.000	15:02
03 *Subject Test	RFI*	15:03
04 Room Air	0.000	15:04

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Gas	0.076	15:08
03 Room Air	0.000	15:08
04 Std. Gas	0.078	15:09
05 Room Air	0.000	15:09
06 Std. Gas	0.077	15:10
07 Room Air	0.000	15:10

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 15:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:20
02 Std. Gas	0.078	15:20
03 Room Air	0.000	15:21
04 Std. Gas	0.078	15:21
05 Room Air	0.000	15:22
06 Std. Gas	0.078	15:22
07 Room Air	0.000	15:23

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-b*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
06/14/2022 15:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:26
02 Std. Gas	0.078	15:26
03 Room Air	0.000	15:27
04 Std. Gas	0.079	15:27
05 Room Air	0.000	15:28
06 Std. Gas	0.079	15:28
07 Room Air	0.000	15:29

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #7-9*