

Intoxilyzer® 8000 Serial Number: 80-000506

Location: Tuxl

A. General Setup and Checks:

- 1. Diagnostic Tests Pass and Instrument in Ready Mode
- 2. Breath tube heated
- 3. Date, time and location code (Level 2,E). Re-set if necessary.
- 4. Print test (Level 1,P). Sign and attach test record.
- 5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 916 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
 - 1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3066 Lot #: 202201A Exp. Date: 1/11/24
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 6.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

- 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1E8th156 Exp. Date: NA
 Display reads "Interferent Detect"
- 3. RFI Check (CMS Mode) Display reads "RFI Detect"
- 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02624080A1 Cyl #: 004 Exp. Date: 3/5/23

Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual Inspection.

Replaced breath tube o-ring + simulator return o-ring.
Replaced o-ring on gas cylinder's regulator.
* Note initial dry calibration was low. Instrument was purged
and calibrations met QA etc.

Instrument is acceptable to be used in the field.

Janelle Postchiller
Breath Analyst Signature

6/15/22
Date


Reviewed by

06/15/22
Date

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 11
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 08/20/2015
Slope 657
Intercept -627881

IR Calibration Date: 08/20/2015

	3um	9um
0th Coef(*100):	-22470	-18919
1st Coef(*100):	266505	136109
2nd Coef(*100):	3702	1644
H2O adj(mg/l*10k):	709	510

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *print test*

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Sol.	0.020	13:39
03 Room Air	0.000	13:39
04 Std. Sol.	0.021	13:40
05 Room Air	0.000	13:41
06 Std. Sol.	0.021	13:41
07 Room Air	0.000	13:42

08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 202201A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Low AC Check 0.020

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Sol.	0.299	13:45
03 Room Air	0.000	13:45
04 Std. Sol.	0.300	13:46
05 Room Air	0.000	13:46
06 Std. Sol.	0.302	13:47
07 Room Air	0.000	13:48

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.306*


06/13/2022

13:59

Test	AC	Time
01 Room Air	0.000	14:00
02 *Subject Test	INT*	14:00
03 Room Air	0.000	14:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent test.*

06/13/2022

14:02

Test	AC	Time
01 Diagnostic	OK	14:03
02 Room Air	0.000	14:03
03 *Subject Test	RFI*	14:04
04 Room Air	0.000	14:04

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*

Form 106-I8000

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.066	14:06
03 Room Air	0.000	14:06
04 Std. Gas	0.068	14:07
05 Room Air	0.000	14:07
06 Std. Gas	0.070	14:08
07 Room Air	0.000	14:08

) Low

Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks:

Dry Calibration Check

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:07
02 Std. Gas	0.081	08:07
03 Room Air	0.000	08:08
04 Std. Gas	0.081	08:08
05 Room Air	0.000	08:09
06 Std. Gas	0.081	08:09
07 Room Air	0.000	08:10


Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Operator Signature
JANELLE PORTSCHELLER

Remarks:

Dry Calibration Check #1-3

06/15/2022

08:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:11
02 Std. Gas	0.080	08:11
03 Room Air	0.000	08:12
04 Std. Gas	0.080	08:12
05 Room Air	0.000	08:13
06 Std. Gas	0.081	08:13
07 Room Air	0.000	08:14

Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*

Form 106-I8000

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:17
02 Std. Gas	0.080	08:18
03 Room Air	0.000	08:18
04 Std. Gas	0.081	08:19
05 Room Air	0.000	08:19
06 Std. Gas	0.081	08:20
07 Room Air	0.000	08:20

Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*

