

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 86-004983

Location: TUCL

A. General Setup and Checks:

1. ☒ Diagnostic Tests Pass and Instrument in Ready Mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 964 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

☒ Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. ☒ Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
- b. ☒ High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

☒ Configure dry gas standard for the following tests (Level 1,S).

2. ☒ Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1C56 Exp. Date: NA
☒ Display reads "Interferent Detect"
3. ☒ RFI Check (CMS Mode) ☒ Display reads "RFI Detect"
4. ☒ Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual Inspection

Replaced Breath tube + Simulator return o-ring.
Replaced external battery

☒ Instrument is acceptable to be used in the field.

Janelle Portscheller
Breath Analyst Signature

6/15/22
Date

[Signature]
Reviewed by

06/16/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 008
Standard Expiration? 10/05/2023
Oper No? 132128

Flow Cal. Date: 04/20/2018
Slope 609
Intercept -462839

IR Calibration Date: 04/20/2018
 3um 9um

0th Coef(*100):	-22916	-27195
1st Coef(*100):	261768	134580
2nd Coef(*100):	1749	1216
H2O adj(mg/l*10k):	496	540

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:32
02 Std. Sol.	0.018	12:33
03 Room Air	0.000	12:34
04 Std. Sol.	0.021	12:34
05 Room Air	0.000	12:35
06 Std. Sol.	0.018	12:36
07 Room Air	0.000	12:36


08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 202201A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Low Ac Check 0.020

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:38
02 Std. Sol.	0.299	12:39
03 Room Air	0.000	12:40
04 Std. Sol.	0.302	12:41
05 Room Air	0.000	12:41
06 Std. Sol.	0.301	12:42
07 Room Air	0.000	12:42

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

High AC check 0.300

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:43

Test	AC	Time
01 Room Air	0.000	12:44
02 *Subject Test	INT*	12:45
03 Room Air	0.000	12:45

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:46

Test	AC	Time
01 Diagnostic	OK	12:46
02 Room Air	0.000	12:47
03 *Subject Test	RFI*	12:47
04 Room Air	0.000	12:48

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:49
02 Std. Gas	0.080	12:50
03 Room Air	0.000	12:50
04 Std. Gas	0.080	12:50
05 Room Air	0.000	12:51
06 Std. Gas	0.080	12:51
07 Room Air	0.000	12:52

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:53
02 Std. Gas	0.080	12:53
03 Room Air	0.000	12:54
04 Std. Gas	0.081	12:54
05 Room Air	0.000	12:55
06 Std. Gas	0.081	12:55
07 Room Air	0.000	12:56


Lot No = 02621080A1

Cyl No = 31

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.14.00 09/16
06/15/2022 12:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Gas	0.081	12:58
03 Room Air	0.000	12:58
04 Std. Gas	0.082	12:59
05 Room Air	0.000	12:59
06 Std. Gas	0.081	12:59
07 Room Air	0.000	13:00

Lot No = 02621080A1

Cyl No = 31

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #7-9*