

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006499

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 997 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
MP6036 ← Sim. Ser #: 2134090 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1C56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0262102A1 Cyl #: 002 Exp. Date: 3/5/23

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average <u>0.081</u>					

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring. Replaced external battery.
Added electrical tape to breath tube hose to prevent wear

Instrument is acceptable to be used in the field.

Janelle Pertschler
Breath Analyst Signature

6/15/22
Date

[Signature]
Reviewed by

06/16/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/14/2022 14:40

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%\$^&* ()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%\$^&* ()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 10
Standard Expiration? 10/05/2023
Oper No? 132128

Flow Cal. Date: 08/18/2015
Slope 678
Intercept -685002

IR Calibration Date: 08/18/2015
3um 9um

0th Coef(*100): -22123 -16014
1st Coef(*100): 264933 138505
2nd Coef(*100): 3281 1621
H2O adj(mg/l*10k): 704 456

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 08:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:26
02 Std. Sol.	0.020	08:29
03 Room Air	0.000	08:30
04 Std. Sol.	0.020	08:31
05 Room Air	0.000	08:31
06 Std. Sol.	0.019	08:32
07 Room Air	0.000	08:33

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 08:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:39
02 Std. Sol.	0.301	08:39
03 Room Air	0.000	08:40
04 Std. Sol.	0.300	08:41
05 Room Air	0.000	08:42
06 Std. Sol.	0.300	08:42
07 Room Air	0.000	08:43

08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 202201F

County = 08

Oper No. = 132128

Janelle Portscher

Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 08:47

Test	AC	Time
01 Room Air	0.000	08:48
02 *Subject Test	INT*	08:48
03 Room Air	0.000	08:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 08:54

Test	AC	Time
01 Diagnostic	OK	08:55
02 Room Air	0.000	08:55
03 *Subject Test	RFI*	08:56
04 Room Air	0.000	08:57

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 08:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:59
02 Std. Gas	0.081	08:59
03 Room Air	0.000	09:00
04 Std. Gas	0.081	09:00
05 Room Air	0.000	09:01
06 Std. Gas	0.081	09:01
07 Room Air	0.000	09:02

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 09:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:03
02 Std. Gas	0.081	09:04
03 Room Air	0.000	09:04
04 Std. Gas	0.081	09:05
05 Room Air	0.000	09:05
06 Std. Gas	0.081	09:06
07 Room Air	0.000	09:06

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = TOXL 8164.14.00 09/16
06/15/2022 09:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:08
02 Std. Gas	0.081	09:08
03 Room Air	0.000	09:09
04 Std. Gas	0.081	09:09
05 Room Air	0.000	09:10
06 Std. Gas	0.081	09:10
07 Room Air	0.000	09:11


Lot No = 02621080A1

Cyl No = 2

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*