



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>30-004953</i>	Instrument Location <i>HACTC Pierce County</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *NA per Annual inspection per 06/16/22*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Darren Hedrick</i>	Date <i>6/15/22</i>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Putschlter</i>	Certified Date <i>6/16/22</i>
--	----------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = RUGB      8164.14.00 09/16  
06/15/2022      16:48

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      26021080A1  
Standard Cyl #?      8  
Standard Expiration?      10/05/2023  
Oper No?      130392

Flow Cal. Date:      04/20/2018  
Slope      609  
Intercept      -462839

IR Calibration Date:      04/20/2018

	3um	9um
--	-----	-----

0th Coef(*100):	-22916	-27195
1st Coef(*100):	261768	134580
2nd Coef(*100):	1749	1216
H2O adj(mg/l*10k):	496	540

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DARREN HEIDBREDER

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = RUGB      8164.14.00 09/16  
06/15/2022      17:01

Test	AC	Time
01 Room Air	0.000	17:03
02 *Subject Test	RFI*	17:05
03 Room Air	0.000	17:06

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, TEST T  
Sub DOB = 08/10/1975  
Sub Sex = Male      Weight = 220  
Test = DUI      Cit = 1  
Dr. Lic. = ND/12345566789  
Lot No = 26021080A1  
Cyl No = 8  
Expiration Date = 10/05/2023  
County = 35      Oper No. = 130392



Operator Signature  
DARREN HEIDBREDER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = RUGB      8164.14.00 09/16  
06/15/2022      16:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:57
02 Std. Gas	0.081	16:58
03 Room Air	0.000	16:58
04 Std. Gas	0.081	16:59
05 Room Air	0.000	16:59
06 Std. Gas	0.081	17:00
07 Room Air	0.000	17:00

Lot No = 26021080A1

Cyl No = 8

Exp Date = 10/05/2023

County = 35

Oper No. = 130392



Operator Signature  
DARREN HEIDBREDER

Remarks:

Form 106-18000