

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004199

Location: TDXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 214 psi Regulator: 250 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 6.020 AC
Sim. Ser #: MP3060 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 20229F Exp. Date: 1/18/24
 - Configure dry gas standard for the following tests (Level 1,S).
 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
 3. RFI Check (CMS Mode) Display reads "RFI Detect"
 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02021080A1 Cyl #: 2 Exp. Date: 3/5/23

| | | | | | |
|---------|--------------|--------|--------------|--------|--------------|
| Test 1 | <u>0.078</u> | Test 4 | <u>0.078</u> | Test 7 | <u>0.079</u> |
| Test 2 | <u>0.078</u> | Test 5 | <u>0.078</u> | Test 8 | <u>0.077</u> |
| Test 3 | <u>0.079</u> | Test 6 | <u>0.078</u> | Test 9 | <u>0.077</u> |
| Average | <u>0.078</u> | | | | |

C. Remarks/Maintenance: Annual inspection

Replaced "B" Replaced breath tube o-ring + simulator return o-ring.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Postchiller
Breath Analyst Signature

6/16/22
Date

[Signature]
Reviewed by

062322
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 11:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 26
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 06/17/2019
Slope 593
Intercept -481610

IR Calibration Date: 06/17/2019
 3um 9um

0th Coef(*100): -17512 -19237
1st Coef(*100): 277207 136841
2nd Coef(*100): 1620 1027
H2O adj(mg/l*10k): 211 375

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print + kept*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 11:47

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:47 |
| 02 Std. Sol. | 0.020 | 11:48 |
| 03 Room Air | 0.000 | 11:48 |
| 04 Std. Sol. | 0.020 | 11:49 |
| 05 Room Air | 0.000 | 11:50 |
| 06 Std. Sol. | 0.020 | 11:50 |
| 07 Room Air | 0.000 | 11:51 |

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 11:57

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:58 |
| 02 Std. Sol. | 0.302 | 11:58 |
| 03 Room Air | 0.000 | 11:59 |
| 04 Std. Sol. | 0.302 | 12:00 |
| 05 Room Air | 0.000 | 12:00 |
| 06 Std. Sol. | 0.303 | 12:01 |
| 07 Room Air | 0.000 | 12:01 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 202201F

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 12:18

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 12:19 |
| 02 *Subject Test | INT* | 12:19 |
| 03 Room Air | 0.000 | 12:20 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 12:27

| Test | AC | Time |
|------------------|-------|-------|
| 01 Diagnostic | OK | 12:28 |
| 02 Room Air | 0.000 | 12:29 |
| 03 *Subject Test | RFI* | 12:31 |
| 04 Room Air | 0.000 | 12:32 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 12:35

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 12:35 |
| 02 Std. Gas | 0.078 | 12:36 |
| 03 Room Air | 0.000 | 12:36 |
| 04 Std. Gas | 0.078 | 12:36 |
| 05 Room Air | 0.000 | 12:37 |
| 06 Std. Gas | 0.079 | 12:37 |
| 07 Room Air | 0.000 | 12:38 |

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 12:42

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 12:43 |
| 02 Std. Gas | 0.078 | 12:43 |
| 03 Room Air | 0.000 | 12:44 |
| 04 Std. Gas | 0.078 | 12:44 |
| 05 Room Air | 0.000 | 12:44 |
| 06 Std. Gas | 0.078 | 12:45 |
| 07 Room Air | 0.000 | 12:45 |

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.14.00 09/16
06/16/2022 12:55

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 12:55 |
| 02 Std. Gas | 0.079 | 12:56 |
| 03 Room Air | 0.000 | 12:56 |
| 04 Std. Gas | 0.077 | 12:56 |
| 05 Room Air | 0.000 | 12:57 |
| 06 Std. Gas | 0.077 | 12:57 |
| 07 Room Air | 0.000 | 12:58 |

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*