

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-005944

Location: TUCL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 486 psi Regulator: 500 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3067 Lot #: 21426 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 4 Exp. Date: 3/5/23

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection

Replaced breath tube o-ring and simulator return o-ring.

Instrument is acceptable to be used in the field.

Janell Pertschiller  
Breath Analyst Signature

6/16/22  
Date

[Signature]  
Reviewed by

061622  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005944  
Location = TOXL                    8164.14.00 09/16  
06/16/2022                            11:18

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            02621080A1  
Standard Cyl #?            27  
Standard Expiration?       03/05/2023  
Oper No?                    132128

Flow Cal. Date:            06/20/2013  
Slope                        677  
Intercept                   -702620

IR Calibration Date:       06/09/2017  
                                  3um                    9um

	3um	9um
0th Coef(*100):	-16158	-17815
1st Coef(*100):	266780	136962
2nd Coef(*100):	1117	1089
H2O adj(mg/l*10k):	317	364

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005944  
Location = TOXL                    8164.14.00 09/16  
06/16/2022                            11:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:40
02 Std. Sol.	0.019	11:41
03 Room Air	0.000	11:42
04 Std. Sol.	0.020	11:42
05 Room Air	0.000	11:43
06 Std. Sol.	0.020	11:44
07 Room Air	0.000	11:44

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*Low AC check 0.020 JJP 06/16/22*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005944  
Location = TOXL      8164.14.00 09/16  
06/16/2022      11:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Sol.	0.302	11:56
03 Room Air	0.000	11:57
04 Std. Sol.	0.303	11:58
05 Room Air	0.000	11:58
06 Std. Sol.	0.303	11:59
07 Room Air	0.000	12:00

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High High AC check 0.300*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005944  
Location = TOXL                    8164.14.00 09/16  
06/16/2022                            12:10

Test	AC	Time
01 Room Air	0.000	12:11
02 *Subject Test	INT*	12:11
03 Room Air	0.000	12:12

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = DUI                                Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 4  
Expiration Date = 03/05/2023  
County = 08                                Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005944  
Location = TOXL                    8164.14.00 09/16  
06/16/2022                            12:13

Test	AC	Time
01 Diagnostic	OK	12:14
02 Room Air	0.000	12:14
03 *Subject Test	RFI*	12:15
04 Room Air	0.000	12:15

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = DUI                                Cit = RFI CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 4  
Expiration Date = 03/05/2023  
County = 08                                Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005944  
Location = TOXL                 8164.14.00 09/16  
06/16/2022                        12:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:16
02 Std. Gas	0.081	12:16
03 Room Air	0.000	12:17
04 Std. Gas	0.081	12:17
05 Room Air	0.000	12:18
06 Std. Gas	0.081	12:18
07 Room Air	0.000	12:19

Lot No = 02621080A1  
Cyl No = 4  
Exp Date = 03/05/2023  
County = 08                        Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005944  
Location = TOXL      8164.14.00 09/16  
06/16/2022      12:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:20
02 Std. Gas	0.081	12:21
03 Room Air	0.000	12:21
04 Std. Gas	0.081	12:22
05 Room Air	0.000	12:22
06 Std. Gas	0.082	12:22
07 Room Air	0.000	12:23

Lot No = 02621080A1  
Cyl No = 4  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-914-6*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-005944  
Location = TOXL 8164.14.00 09/16  
06/16/2022 12:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:29
02 Std. Gas	0.082	12:30
03 Room Air	0.000	12:30
04 Std. Gas	0.082	12:31
05 Room Air	0.000	12:31
06 Std. Gas	0.081	12:32
07 Room Air	0.000	12:32


Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*

Form 106-I8000