

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006490 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 899 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3070 Lot #: 22CP 41184 Exp. Date: 3/1/24
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.400 AC  
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3247 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02624080A1 Cyl #: 016 Exp. Date: 3/5/23

Test 1 <u>0.078</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection.  
Replaced breath tube oring. Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Portschiller  
Breath Analyst Signature

6/16/22  
Date

[Signature]  
Reviewed by

06/16/22  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-006490  
Location = ASHL                8164.14.00 09/16  
06/16/2022                       12:28

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?         No  
Display Volume?             No  
Disable On Memfull?         Yes  
# of Print Copies?           1  
Select Std (D/W/I)?         Dry  
Standard Value?             0.080  
Standard Lot #?             02621080A1  
Standard Cyl #?             36  
Standard Expiration?        03/05/2023  
Oper No?                    132128

Flow Cal. Date:             09/04/2020  
Slope                        649  
Intercept                    -598817

IR Calibration Date:        09/04/2020  
                                  3um                    9um

-----  
0th Coef(\*100):            -25190                -16392  
1st Coef(\*100):            271645                136906  
2nd Coef(\*100):            2533                    1222  
H2O adj(mg/l\*10k):        460                    309

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = ASHL      8164.14.00 09/16  
06/16/2022      12:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:33
02 Std. Sol.	0.020	12:34
03 Room Air	0.000	12:35
04 Std. Sol.	0.020	12:35
05 Room Air	0.000	12:36
06 Std. Sol.	0.020	12:37
07 Room Air	0.000	12:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3070  
Std Sol No = 22CP41184  
County = 08      Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check*      *20 July 22*  
*0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = ASHL      8164.14.00 09/16  
06/16/2022      12:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:41
02 Std. Sol.	0.395	12:43
03 Room Air	0.000	12:43
04 Std. Sol.	0.396	12:44
05 Room Air	0.000	12:45
06 Std. Sol.	0.397	12:45
07 Room Air	0.000	12:46

08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC Check 0.400*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = ASHL      8164.14.00 09/16  
06/16/2022      12:48

Test	AC	Time
01 Room Air	0.000	12:49
02 *Subject Test	INT*	12:49
03 Room Air	0.000	12:50

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 16  
Expiration Date = 03/05/2023  
County = 08      Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006490  
Location = ASHL                    8164.14.00 09/16  
06/16/2022                            12:50

Test	AC	Time
01 Diagnostic	OK	12:51
02 Room Air	0.000	12:52
03 *Subject Test	RFI*	12:52
04 Room Air	0.000	12:53

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS921456  
Lot No = 02621080A1  
Cyl No = 16  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: ~~Dry Calibration~~ *RFI check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006490  
Location = ASHL                    8164.14.00 09/16  
06/16/2022                            12:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:54
02 Std. Gas	0.078	12:54
03 Room Air	0.000	12:55
04 Std. Gas	0.079	12:55
05 Room Air	0.000	12:56
06 Std. Gas	0.079	12:56
07 Room Air	0.000	12:57

Lot No = 02621080A1  
Cyl No = 16  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006490  
Location = ASHL      8164.14.00 09/16  
06/16/2022      12:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Gas	0.080	12:59
03 Room Air	0.000	13:00
04 Std. Gas	0.079	13:00
05 Room Air	0.000	13:01
06 Std. Gas	0.079	13:01
07 Room Air	0.000	13:02

Lot No = 02621080A1  
Cyl No = 16  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006490  
Location = ASHL                    8164.14.00 09/16  
06/16/2022                            13:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:06
02 Std. Gas	0.079	13:06
03 Room Air	0.000	13:06
04 Std. Gas	0.079	13:07
05 Room Air	0.000	13:07
06 Std. Gas	0.079	13:08
07 Room Air	0.000	13:08

Lot No = 02621080A1  
Cyl No = 16  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*