



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number 80-006505	Instrument Location Renville County Sheriff's Office
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Daniel Weiss</i>	Date 06 17 2022
--	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Alan Meul</i>	Certified Date 09 Feb 2023
---	-------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006505  
Location = MOHL                    8164.14.00 09/16  
06/17/2022                            17:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:54
02 Std. Gas	0.080	17:54
03 Room Air	0.000	17:55
04 Std. Gas	0.080	17:55
05 Room Air	0.000	17:56
06 Std. Gas	0.079	17:56
07 Room Air	0.000	17:57

Lot No = 26021080A1  
Cyl No = 23  
Exp Date = 10/05/2023  
County = 38                            Oper No. = 131142



Operator Signature  
DANIEL MOSS

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006505  
Location = MOHL                    8164.14.00 09/16  
06/17/2022                            17:57

Test	AC	Time
01 Room Air	0.000	17:59
02 *Subject Test	RFI*	17:59
03 Room Air	0.000	17:59

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST NA  
Sub DOB = 07/05/1999  
Sub Sex = Male                            Weight = 230  
Test = DUI    Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 26021080A1  
Cyl No = 23  
Expiration Date = 10/05/2023  
County = 38                                    Oper No. = 131142



Operator Signature  
DANIEL MOSS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006505  
Location = MOHL                8164.14.00 09/16  
06/17/2022                      17:47

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            26021080A1  
Standard Cyl #?            23  
Standard Expiration?       10/05/2023  
Oper No?                    131142

Flow Cal. Date:            08/20/2015  
Slope                      671  
Intercept                  -642269

IR Calibration Date:        08/20/2015  
                                  3um                    9um

-----  
0th Coef(\*100):            -29025                -17051  
1st Coef(\*100):            252705                133345  
2nd Coef(\*100):            3840                   1546  
H2O adj(mg/l\*10k):        817                    461

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DANIEL MOSS

Remarks: