



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL


CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number 80-004945	Instrument Location BMDC
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) Annual Check Completed	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): None Needed
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 6/17/2022
--	-------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 6/20/22
---	---------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004945  
Location = BMDC                8164.14.00 09/16  
06/17/2022                      15:10

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            02621080A1  
Standard Cyl #?            7  
Standard Expiration?       03/05/2023  
Oper No?                    130315

Flow Cal. Date:            09/26/2012  
Slope                      622  
Intercept                  -485836

IR Calibration Date:       04/22/2016  
                                    3um                      9um

-----  
0th Coef(\*100):            -23427                      -26406  
1st Coef(\*100):            266989                      144341  
2nd Coef(\*100):            2085                        1078  
H2O adj(mg/l\*10k):        539                         400

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
JARED LEMIEUX

Remarks: *Print Test after install  
(Annual check-up) GOOD*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = BMDC      8164.14.00 09/16  
06/17/2022      15:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:17
02 Std. Gas	0.079	15:18
03 Room Air	0.000	15:18
04 Std. Gas	0.079	15:19
05 Room Air	0.000	15:19
06 Std. Gas	0.078	15:19
07 Room Air	0.000	15:20

Lot No = 02621080A1  
Cyl No = 7  
Exp Date = 03/05/2023  
County = 08      Oper No. = 130315

  
Operator Signature  
JARED LEMIEUX

Remarks: Calibration Test after install  
(Annual check-up) GOOD


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = BMDC      8164.14.00 09/16  
06/17/2022      15:12

Test	AC	Time
01 Room Air	0.000	15:14
02 *Subject Test	RFI*	15:14
03 Room Air	RFI*	15:14

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, INSTALL NA  
Sub DOB = 01/01/1990  
Sub Sex = Unknown      Weight = NA  
Test = DUI      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 02621080A1  
Cyl No = 7  
Expiration Date = 03/05/2023  
County = 08      Oper No. = 130315

  
Operator Signature  
JARED LEMIEUX

Remarks:

*RFI check after install  
(Annual check-up)*

Form 106-I8000

GOOD