

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-604944 Location: TKL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 785 psi Regulator: 800 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
MP5290 ← Sim. Ser #: 2134080 Lot #: 21340 Exp. Date: 8/24/23
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 031 Exp. Date: 3/5/23

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.079</u>	Test 9	<u>0.080</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: Annual inspection.
Replaced breath tube and simulator return o-rings
Replaced the electrical tape that was on the breath tube hose.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

6/22/22
Date

Reviewed by _____

062872
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 17
Standard Expiration? 10/05/2023
Oper No? 132128

Flow Cal. Date: 07/01/2011
Slope 685
Intercept -634740

IR Calibration Date: 07/01/2011

	3um	9um
0th Coef(*100):	-22591	-17234
1st Coef(*100):	265915	132691
2nd Coef(*100):	2841	1562
H2O adj(mg/l*10k):	559	425

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Sol.	0.019	10:07
03 Room Air	0.000	10:07
04 Std. Sol.	0.019	10:08
05 Room Air	0.000	10:08
06 Std. Sol.	0.020	10:09
07 Room Air	0.000	10:10

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:12
02 Std. Sol.	0.304	10:13
03 Room Air	0.000	10:13
04 Std. Sol.	0.305	10:14
05 Room Air	0.000	10:15
06 Std. Sol.	0.305	10:15
07 Room Air	0.000	10:16

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 21420
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *End of High AC Check 0.300*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:19

Test	AC	Time
01 Room Air	0.000	10:19
02 *Subject Test	INT*	10:20
03 Room Air	0.000	10:21

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:22

Test	AC	Time
01 Diagnostic	OK	10:23
02 Room Air	0.000	10:24
03 *Subject Test	RFI*	10:24
04 Room Air	0.000	10:24

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.079	10:36
03 Room Air	0.000	10:37
04 Std. Gas	0.079	10:37
05 Room Air	0.000	10:38
06 Std. Gas	0.080	10:38
07 Room Air	0.000	10:38

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:40
02 Std. Gas	0.079	10:40
03 Room Air	0.000	10:40
04 Std. Gas	0.079	10:41
05 Room Air	0.000	10:41
06 Std. Gas	0.079	10:42
07 Room Air	0.000	10:42

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.14.00 09/16
06/22/2022 10:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Gas	0.080	10:43
03 Room Air	0.000	10:44
04 Std. Gas	0.080	10:44
05 Room Air	0.000	10:45
06 Std. Gas	0.080	10:45
07 Room Air	0.000	10:46

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*