

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004947

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 958 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.08 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 49 Exp. Date: 2/5/24

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.082</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.082</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube and simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

6/22/22
Date

[Signature]
Reviewed by

062322
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 13:19

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Sol.	0.018	13:21
03 Room Air	0.000	13:21
04 Std. Sol.	0.019	13:22
05 Room Air	0.000	13:22
06 Std. Sol.	0.018	13:23
07 Room Air	0.000	13:24

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 13:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:25
02 Std. Sol.	0.294	13:27
03 Room Air	0.000	13:27
04 Std. Sol.	0.295	13:28
05 Room Air	0.000	13:29
06 Std. Sol.	0.295	13:29
07 Room Air	0.000	13:30


08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 13:36

Test	AC	Time
01 Room Air	0.000	13:37
02 *Subject Test	INT*	13:38
03 Room Air	0.000	13:38

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CHK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 13:41

Test	AC	Time
01 Diagnostic	OK	13:42
02 Room Air	0.000	13:42
03 *Subject Test	RFI*	13:44
04 Room Air	0.000	13:44

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 13:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:58
02 Std. Gas	0.081	13:58
03 Room Air	0.000	13:59
04 Std. Gas	0.082	13:59
05 Room Air	0.000	14:00
06 Std. Gas	0.082	14:00
07 Room Air	0.000	14:00

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Gas	0.081	14:02
03 Room Air	0.000	14:02
04 Std. Gas	0.081	14:03
05 Room Air	0.000	14:03
06 Std. Gas	0.080	14:03
07 Room Air	0.000	14:04

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = TOXL 8164.14.00 09/16
06/22/2022 14:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.081	14:06
03 Room Air	0.000	14:06
04 Std. Gas	0.081	14:06
05 Room Air	0.000	14:07
06 Std. Gas	0.081	14:07
07 Room Air	0.000	14:08

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*