



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Luke C. Olsen</i>		
Location <i>Killdeer PD</i>	Intoxilyzer® Serial Number <i>80-004938</i>	
Gas Lot Number <i>0262108641</i>	Gas Cylinder Number <i>005</i>	Gas Expiration Date <i>03/05/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
gas 6/23/22 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.076 AC *gas 6/23/22*
 - B. 0.076 AC *gas 6/23/22*
 - C. 0.076 AC *gas 6/23/22*
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Luke C. Olsen</i>	Date <i>06/22/2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Pertscheller</i>	Date <i>23 June 22</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = KDPD 8164.14.00 09/16
06/22/2022 04:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:42
02 Std. Gas	0.076	04:42
03 Room Air	0.000	04:43
04 Std. Gas	0.076	04:43
05 Room Air	0.000	04:44
06 Std. Gas	0.076	04:44
07 Room Air	0.000	04:45

Lot No = 02621080A1
Cyl No = 5
Exp Date = 03/05/2023
County = 13 Oper No. = 131669


Operator Signature
LUKE OLSON

Remarks: