

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-005362 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 888 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: 1656 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34921080AI Cyl #: 47 Exp. Date: 2/5/24

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.079</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.081</u>
Average	<u>0.0798</u> <u>0.080</u>				

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube and simulator return o-rings. Replaced electrical taps on breath hose to prevent wear.

Instrument is acceptable to be used in the field.

Janelle Pertschauer  
Breath Analyst Signature

6/23/22  
Date

[Signature]  
Reviewed by

062822  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = TOXL                 8164.14.00 09/16  
06/23/2022                        09:37

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 33321080A4  
Standard Cyl #?                 12  
Standard Expiration?            01/05/2024  
Oper No?                        132128  
  
Flow Cal. Date:                 05/22/2020  
Slope                            659  
Intercept                       -620990

IR Calibration Date:            05/26/2020  
                                    3um                            9um

-----  
0th Coef(\*100):                -20883                        -25370  
1st Coef(\*100):                265508                        144960  
2nd Coef(\*100):                936                            714  
H2O adj(mg/l\*10k):            391                            379

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.14.00 09/16  
06/23/2022      09:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Sol.	0.018	09:47
03 Room Air	0.000	09:47
04 Std. Sol.	0.018	09:48
05 Room Air	0.000	09:48
06 Std. Sol.	0.018	09:49
07 Room Air	0.000	09:49


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 21340

County = 08

Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            09:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:52
02 Std. Sol.	0.292	10:03
03 Room Air	0.000	10:03
04 Std. Sol.	0.294	10:04
05 Room Air	0.000	10:05
06 Std. Sol.	0.293	10:05
07 Room Air	0.000	10:06

08 Sim Temp = 34.0°C

Simul Ser No = MP3068  
Std Sol No = 202201F  
County = 08                            Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            10:08

Test	AC	Time
01 Room Air	0.000	10:09
02 *Subject Test	INT*	10:10
03 Room Air	0.000	10:10

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CHK  
Dr. Lic. = ND/DIS921456  
Lot No = 34921080A1  
Cyl No = 47  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = TOXL                 8164.14.00 09/16  
06/23/2022                        10:11

Test	AC	Time
01 Diagnostic	OK	10:12
02 Room Air	0.000	10:12
03 *Subject Test	RFI*	10:14
04 Room Air	0.000	10:15

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS921456  
Lot No = 34921080A1  
Cyl No = 47  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = TOXL                    8164.14.00 09/16  
06/23/2022                            10:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:17
02 Std. Gas	0.080	10:18
03 Room Air	0.000	10:18
04 Std. Gas	0.081	10:19
05 Room Air	0.000	10:19
06 Std. Gas	0.079	10:19
07 Room Air	0.000	10:20

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.14.00 09/16  
06/23/2022      10:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:27
02 Std. Gas	0.080	10:27
03 Room Air	0.000	10:28
04 Std. Gas	0.080	10:28
05 Room Air	0.000	10:29
06 Std. Gas	0.079	10:29
07 Room Air	0.000	10:30

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration checks #4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.14.00 09/16  
06/23/2022      10:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Gas	0.079	10:33
03 Room Air	0.000	10:33
04 Std. Gas	0.080	10:33
05 Room Air	0.000	10:34
06 Std. Gas	0.081	10:34
07 Room Air	0.000	10:35

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Checks # 7-9*