

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 86-004189

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 978 psi Regulator: 975 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3547 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.08 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 47 Exp. Date: 2/5/24

Test 1	<u>0.080</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.079</u>	Test 6	<u>0.080</u>	Test 9	<u>0.079</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring and simulator return o-ring.

Instrument is acceptable to be used in the field.

Janelle Protschuller
Breath Analyst Signature

6/23/22
Date

[Signature]
Reviewed by

062822
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 11:39

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 48
Standard Expiration? 04/05/2022
Oper No? 132128

Flow Cal. Date: 02/02/2011
Slope 628
Intercept -452608

IR Calibration Date: 05/19/2016
 3um 9um

0th Coef(*100): -17767 -18761
1st Coef(*100): 263439 133606
2nd Coef(*100): 1960 1140
H2O adj(mg/l*10k): 341 362

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 11:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Sol.	0.019	11:42
03 Room Air	0.000	11:43
04 Std. Sol.	0.019	11:44
05 Room Air	0.000	11:44
06 Std. Sol.	0.020	11:45
07 Room Air	0.000	11:45


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 11:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Sol.	0.295	11:48
03 Room Air	0.000	11:48
04 Std. Sol.	0.298	11:49
05 Room Air	0.000	11:50
06 Std. Sol.	0.297	11:50
07 Room Air	0.000	11:51

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 202201F
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: High AC check ~~0.000~~ 0.300


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 11:55

Test	AC	Time
01 Room Air	0.000	11:56
02 *Subject Test	INT*	11:57
03 Room Air	0.000	11:57

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CHK
Dr. Lic. = ND/DIS921456
Lot No = 34921080A1
Cyl No = 47
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 11:58

Test	AC	Time
01 Diagnostic	OK	11:59
02 Room Air	0.000	11:59
03 *Subject Test	RFI*	12:00
04 Room Air	0.000	12:00

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 34921080A1
Cyl No = 47
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 12:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:10
02 Std. Gas	0.079	12:10
03 Room Air	0.000	12:10
04 Std. Gas	0.080	12:11
05 Room Air	0.000	12:11
06 Std. Gas	0.080	12:12
07 Room Air	0.000	12:12

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
06/23/2022 12:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:13
02 Std. Gas	0.080	12:13
03 Room Air	0.000	12:14
04 Std. Gas	0.080	12:14
05 Room Air	0.000	12:15
06 Std. Gas	0.079	12:15
07 Room Air	0.000	12:15

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #7-9*