



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Mckenzie County Jail

Serial Number <i>00-005362</i>	Instrument Location <i>1201 12th St. SE, Watford City, ND 58854</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <i>INSPECTION</i>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>6/23/22</i>
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Patschiller</i>	Certified Date <i>06/30/2022</i>
---	-------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = MCSO 8164.14.00 09/16
06/23/2022 14:05

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33321080A4
Standard Cyl #? 12
Standard Expiration? 01/05/2024
Oper No? 132007

Flow Cal. Date: 05/22/2020
Slope 659
Intercept -620990

IR Calibration Date: 05/26/2020
3um 9um

0th Coef(*100): -20883 -25370
1st Coef(*100): 265508 144960
2nd Coef(*100): 936 714
H2O adj(mg/l*10k): 391 379

***** Printer Test End *****



Operator Signature
KARI STUART


STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }^{ss}

Remarks:

Print test

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

Form 106-I8000 on this 22nd day of JUNE 2022


McKenzie County Sheriff's Department

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = MCSO 8164.14.00 09/16
06/23/2022 14:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:16
02 Std. Gas	0.081	14:16
03 Room Air	0.000	14:17
04 Std. Gas	0.080	14:17
05 Room Air	0.000	14:18
06 Std. Gas	0.079	14:18
07 Room Air	0.000	14:19

Lot No = 33321080A4
Cyl No = 12
Exp Date = 01/05/2024
County = 27 Oper No. = 132007

Kari Stuart #3410

Operator Signature
KARI STUART

Remarks:

ACA

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 22nd day of JUNE, 2022

Kari Stuart #3410
Mckenzie County Sheriff's Department

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005362
Location = MCSO 8164.14.00 09/16
06/23/2022 14:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	14:20
02 Room Air	0.000	14:20

*Invalid Test
Inhibited - RFI

Lot No = 33321080A4
Cyl No = 12
Exp Date = 01/05/2024
County = 27 Oper No. = 132007

Kari Stuart #3410

Operator Signature
KARI STUART

Remarks:

RFI

Form 106-I8000

STATE OF NORTH DAKOTA }
COUNTY OF MCKENZIE }

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 27th day of JUNE, 2022
Kari Stuart #3410
McKenzie County Sheriff's Department