



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-004944</b>	Instrument Location <b>WILL</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> #456	Date <b>06-23-2022</b>
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Pertschiller</i>	Certified Date <b>07/04/22</b>
--	-----------------------------------




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004944  
Location = WILL                    8164.14.00 09/16  
06/23/2022                            19:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:29
02 Std. Gas	0.080	19:30
03 Room Air	0.000	19:30
04 Std. Gas	0.079	19:31
05 Room Air	0.000	19:31
06 Std. Gas	0.080	19:32
07 Room Air	0.000	19:32

Lot No = 26021080A1  
Cyl No = 17  
Exp Date = 10/05/2023  
County = 53                            Oper No. = 130755

  
\_\_\_\_\_  
Operator Signature  
JUSTIN PELZL

Remarks:  
*Install  
ACA Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004944  
Location = WILL      8164.14.00 09/16  
06/23/2022      19:34

Test	AC	Time
01 Room Air	RFI*	19:37
02 Room Air	RFI*	19:38

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, RFI TEST INSTALL  
Sub DOB = 01/01/1900  
Sub Sex = Unknown      Weight = NA  
Test = OTH      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 26021080A1  
Cyl No = 17  
Expiration Date = 10/05/2023  
County = 53      Oper No. = 130755

Justin Pelzl #456  
Operator Signature  
JUSTIN PELZL

Remarks:

Install  
RFI Test