

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003061 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 950 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.630 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 207201F Exp. Date: 1/18/24
 - Configure dry gas standard for the following tests (Level 1,S).
 2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
 3. RFI Check (CMS Mode) Display reads "RFI Detect"
 4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 47 Exp. Date: 2/5/24

Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.078</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.078</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection.
Replaced breath tube and simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Pertscheller
Breath Analyst Signature

6/23²⁰ 6/24/22
Date

[Signature]
Reviewed by

062522
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 11:50

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 26
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 03/04/2011
Slope 700
Intercept -623726

IR Calibration Date: 04/10/2015

	3um	9um
0th Coef(*100):	-25722	-17879
1st Coef(*100):	266163	136703
2nd Coef(*100):	2302	1151
H2O adj(mg/l*10k):	412	294

***** Printer Test End *****

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 11:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Sol.	0.018	11:56
03 Room Air	0.000	11:57
04 Std. Sol.	0.018	11:57
05 Room Air	0.000	11:58
06 Std. Sol.	0.018	11:59
07 Room Air	0.000	11:59


08 Sim Temp = 34.0°C

Simul Ser No = MP3036

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low ac Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 12:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Sol.	0.292	12:04
03 Room Air	0.000	12:04
04 Std. Sol.	0.292	12:05
05 Room Air	0.000	12:05
06 Std. Sol.	0.293	12:06
07 Room Air	0.000	12:07

08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 202201F

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 12:09

Test	AC	Time
01 Room Air	0.000	12:10
02 *Subject Test	INT*	12:10
03 Room Air	0.000	12:11

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS921456
Lot No = 34921080A1
Cyl No = 47
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RTI check for
Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 12:23

Test	AC	Time
01 Diagnostic	OK	12:24
02 Room Air	0.000	12:24
03 *Subject Test	RFI*	12:26
04 Room Air	0.000	12:26

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 34921080A1
Cyl No = 47
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 12:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:28
02 Std. Gas	0.078	12:29
03 Room Air	0.000	12:29
04 Std. Gas	0.079	12:29
05 Room Air	0.000	12:30
06 Std. Gas	0.078	12:30
07 Room Air	0.000	12:31

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 12:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:32
02 Std. Gas	0.079	12:32
03 Room Air	0.000	12:33
04 Std. Gas	0.079	12:33
05 Room Air	0.000	12:33
06 Std. Gas	0.079	12:34
07 Room Air	0.000	12:34

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = TOXL 8164.14.00 09/16
06/24/2022 12:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Gas	0.078	12:36
03 Room Air	0.000	12:37
04 Std. Gas	0.079	12:37
05 Room Air	0.000	12:37
06 Std. Gas	0.079	12:38
07 Room Air	0.000	12:38

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128

Janelle Portscheller
Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4^{9/16} 7-9*