

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003060

Location: Tox C

A. General Setup and Checks:

- 1.  Diagnostic Tests Pass and Instrument in Ready Mode
- 2.  Breath tube heated
- 3.  Date, time and location code (Level 2,E). Re-set if necessary.
- 4.  Print test (Level 1,P). Sign and attach test record.
- 5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 143 psi Regulator: 150 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

- 1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
  - b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.400 AC  
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

- 2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DL3847 Lot #: ICS 6 Exp. Date: NA  
 Display reads "Interferent Detect"
- 3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
- 4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 0262108041 Cyl #: 16 Exp. Date: 3/5/23

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.078</u>	Test 8	<u>0.078</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.0782 ≈ 0.079</u>				

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube and simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Putscheller  
Breath Analyst Signature

06/24/2022  
Date

[Signature]  
Reviewed by

062422  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003060  
Location = TOXL                    8164.14.00 09/16  
06/24/2022                            11:51

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    02621080A1  
Standard Cyl #?                    26  
Standard Expiration?                03/05/2023  
Oper No?                            132128

Flow Cal. Date:                    04/08/2008  
Slope                                702  
Intercept                            -647610

IR Calibration Date:                06/04/2008  
                                          3um                                9um

-----  
0th Coef(\*100):                    -5190                                -7737  
1st Coef(\*100):                    270580                                138779  
2nd Coef(\*100):                    2499                                    956  
H2O adj(mg/l\*10k):                377                                    188

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003060  
Location = TOXL                    8164.14.00 09/16  
06/24/2022                            11:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:57
02 Std. Sol.	0.020	11:58
03 Room Air	0.000	11:59
04 Std. Sol.	0.020	11:59
05 Room Air	0.000	12:00
06 Std. Sol.	0.020	12:01
07 Room Air	0.000	12:01

08 Sim Temp = 34.0°C

Simul Ser No = MP3070  
Std Sol No = 22CP41184  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = TOXL      8164.14.00 09/16  
06/24/2022      12:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:04
02 Std. Sol.	0.392	12:07
03 Room Air	0.000	12:08
04 Std. Sol.	0.392	12:08
05 Room Air	0.000	12:09
06 Std. Sol.	0.392	12:10
07 Room Air	0.000	12:10

08 Sim Temp = 34.0°C

Simul Ser No = ~~MP3070~~ *JIP MP3065*  
Std Sol No = 202103E  
County = 08      Oper No. = 132128

*Janelle Portscheller*  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac Check 0.400*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003060  
Location = TOXL                    8164.14.00 09/16  
06/24/2022                            12:12

Test	AC	Time
01 Room Air	0.000	12:13
02 *Subject Test	INT*	12:14
03 Room Air	0.000	12:15

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CHK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 16  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent checked*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003060  
Location = TOXL                    8164.14.00 09/16  
06/24/2022                            12:34

Test	AC	Time
01 Diagnostic	OK	12:38
02 Room Air	0.000	12:38
03 *Subject Test	RFI*	12:39
04 Room Air	0.000	12:40

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 16  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = TOXL      8164.14.00 09/16  
06/24/2022      12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:41
02 Std. Gas	0.079	12:42
03 Room Air	0.000	12:42
04 Std. Gas	0.079	12:42
05 Room Air	0.000	12:43
06 Std. Gas	0.079	12:43
07 Room Air	0.000	12:44

Lot No = 02621080A1  
Cyl No = 16  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003060  
Location = TOXL                    8164.14.00 09/16  
06/24/2022                            12:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:49
02 Std. Gas	0.079	12:49
03 Room Air	0.000	12:50
04 Std. Gas	0.078	12:50
05 Room Air	0.000	12:51
06 Std. Gas	0.079	12:51
07 Room Air	0.000	12:51

Lot No = 02621080A1  
Cyl No = 16  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3 & 4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = TOXL      8164.14.00 09/16  
06/24/2022      12:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:52
02 Std. Gas	0.080	12:53
03 Room Air	0.000	12:53
04 Std. Gas	0.078	12:54
05 Room Air	0.000	12:54
06 Std. Gas	0.079	12:54
07 Room Air	0.000	12:55

Lot No = 02621080A1  
Cyl No = 16  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6 JRP 7-9*